

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1622 SAS No.: Q1622 SDG NO.: Q1622
 Lab File ID: VN086065.D Date Analyzed: 03/24/2025
 Instrument ID: MSVOA_N Time Analyzed: 12:03
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	132834	13.788				
UPPER LIMIT	265668	14.288				
LOWER LIMIT	66417	13.288				
EPA SAMPLE NO.						
MW1	163397	13.79				
MW3	157433	13.79				
Field-Blank	173343	13.79				
VN0324WBL01	85734	13.79				
VN0324WBS01	121424	13.79				
VN0324WBSD01	206640	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.