

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1623 SAS No.: Q1623 SDG NO.: Q1623

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/17/2025 03/17/2025

Client Sample No.: CCAL01 Date Analyzed: 03/25/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070107.D Time Analyzed: 11:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.486	4.386	4.586	230024400.000	228012000.000	-0.9
Aroclor-1016-2	4.504	4.404	4.604	387085800.000	378732000.000	-2.2
Aroclor-1016-3	4.562	4.463	4.663	252199800.000	245008000.000	-2.9
Aroclor-1016-4	4.652	4.551	4.751	199753500.000	195224000.000	-2.3
Aroclor-1016-5	4.936	4.836	5.036	196065600.000	193345000.000	-1.4
Aroclor-1260-1	6.026	5.926	6.126	342744100.000	323296000.000	-5.7
Aroclor-1260-2	6.282	6.181	6.381	375376700.000	380355000.000	1.3
Aroclor-1260-3	6.633	6.533	6.733	301270200.000	294462000.000	-2.3
Aroclor-1260-4	6.847	6.746	6.946	326357800.000	319344000.000	-2.2
Aroclor-1260-5	7.152	7.052	7.252	583624600.000	579075000.000	-0.8
DCB	8.503	8.403	8.603	3797495000.000	3594750000.000	-5.4
TCX	3.405	3.305	3.505	8939766000.000	8777870000.000	-1.8

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1623 SAS No.: Q1623 SDG NO.: Q1623

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/17/2025 03/17/2025

Client Sample No.: CCAL01 Date Analyzed: 03/25/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070107.D Time Analyzed: 11:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.785	3.687	3.887	223585300.000	225508000.000	0.9
Aroclor-1016-2	3.801	3.702	3.902	365003900.000	354482000.000	-2.9
Aroclor-1016-3	3.961	3.863	4.063	206649600.000	202812000.000	-1.9
Aroclor-1016-4	4.009	3.911	4.111	189703100.000	182957000.000	-3.6
Aroclor-1016-5	4.204	4.106	4.306	222979400.000	217659000.000	-2.4
Aroclor-1260-1	5.198	5.100	5.300	400599300.000	382706000.000	-4.5
Aroclor-1260-2	5.387	5.289	5.489	466852700.000	455843000.000	-2.4
Aroclor-1260-3	5.529	5.431	5.631	444827800.000	428714000.000	-3.6
Aroclor-1260-4	5.990	5.892	6.092	392153300.000	383630000.000	-2.2
Aroclor-1260-5	6.234	6.136	6.336	828820000.000	819650000.000	-1.1
DCB	7.545	7.449	7.649	5861421000.000	5760910000.000	-1.7
TCX	2.783	2.683	2.883	8065420000.000	7640250000.000	-5.3