

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1623 SAS No.: Q1623 SDG NO.: Q1623

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/17/2025 03/17/2025

Client Sample No.: CCAL02 Date Analyzed: 03/25/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070115.D Time Analyzed: 16:10

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.485	4.386	4.586	230024400.000	235191000.000	2.2
Aroclor-1016-2	4.504	4.404	4.604	387085800.000	385495000.000	-0.4
Aroclor-1016-3	4.561	4.463	4.663	252199800.000	245453000.000	-2.7
Aroclor-1016-4	4.651	4.551	4.751	199753500.000	198250000.000	-0.8
Aroclor-1016-5	4.936	4.836	5.036	196065600.000	197739000.000	0.9
Aroclor-1260-1	6.025	5.926	6.126	342744100.000	352423000.000	2.8
Aroclor-1260-2	6.281	6.181	6.381	375376700.000	399155000.000	6.3
Aroclor-1260-3	6.632	6.533	6.733	301270200.000	300736000.000	-0.2
Aroclor-1260-4	6.847	6.746	6.946	326357800.000	320665000.000	-1.7
Aroclor-1260-5	7.152	7.052	7.252	583624600.000	598009000.000	2.5
DCB	8.503	8.403	8.603	3797495000.000	3676440000.000	-3.2
TCX	3.405	3.305	3.505	8939766000.000	8830990000.000	-1.2

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1623 SAS No.: Q1623 SDG NO.: Q1623

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/17/2025 03/17/2025

Client Sample No.: CCAL02 Date Analyzed: 03/25/2025

Lab Sample No.: AR1660CCC400 Data File : PQ070115.D Time Analyzed: 16:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.785	3.687	3.887	223585300.000	235522000.000	5.3
Aroclor-1016-2	3.801	3.702	3.902	365003900.000	370384000.000	1.5
Aroclor-1016-3	3.961	3.863	4.063	206649600.000	210383000.000	1.8
Aroclor-1016-4	4.010	3.911	4.111	189703100.000	187835000.000	-1.0
Aroclor-1016-5	4.205	4.106	4.306	222979400.000	229732000.000	3.0
Aroclor-1260-1	5.199	5.100	5.300	400599300.000	391422000.000	-2.3
Aroclor-1260-2	5.388	5.289	5.489	466852700.000	463457000.000	-0.7
Aroclor-1260-3	5.530	5.431	5.631	444827800.000	441611000.000	-0.7
Aroclor-1260-4	5.991	5.892	6.092	392153300.000	389718000.000	-0.6
Aroclor-1260-5	6.235	6.136	6.336	828820000.000	829957000.000	0.1
DCB	7.547	7.449	7.649	5861421000.000	5917270000.000	1.0
TCX	2.783	2.683	2.883	8065420000.000	7787110000.000	-3.5