

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1623 SAS No.: Q1623 SDG NO.: Q1623

EPA Sample No.: SSTD020671 Date Analyzed: 04/01/2025

Lab File ID: BP024205.D Time Analyzed: 09:56

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	472706	7.74	2079510	10.52	1316900	14.38
UPPER LIMIT	945412	8.24	4159020	11.016	2633800	14.881
LOWER LIMIT	236353	7.24	1039760	10.016	658450	13.881
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	281021	7.74	1193690	10.52	806029	14.38

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.