

Surrogate Summary

SDG No.: Q1623

Client: Chemtech Consulting Group

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1623-01	Storage-Blank-SOIL-REF-6	1,2-Dichloroethane-d4	50	52.0	104	74	125
		Dibromofluoromethane	50	49.0	98	75	124
		Toluene-d8	50	49.0	98	86	113
		4-Bromofluorobenzene	50	49.0	98	77	121
Q1623-02	Storage-Blank-WATER-REF-5	1,2-Dichloroethane-d4	50	49.0	98	74	125
		Dibromofluoromethane	50	46.0	92	75	124
		Toluene-d8	50	46.0	92	86	113
		4-Bromofluorobenzene	50	45.0	90	77	121
Q1623-03	Storage-Blank-WATER-REF-4	1,2-Dichloroethane-d4	50	50.0	100	74	125
		Dibromofluoromethane	50	47.0	94	75	124
		Toluene-d8	50	47.0	94	86	113
		4-Bromofluorobenzene	50	47.0	94	77	121
Q1623-04	Storage-Blank-SAMPLE-MANAGEMENT	1,2-Dichloroethane-d4	50	48.0	96	74	125
		Dibromofluoromethane	50	47.0	94	75	124
		Toluene-d8	50	47.0	94	86	113
		4-Bromofluorobenzene	50	46.0	92	77	121
VX0402WBL01	VX0402WBL01	1,2-Dichloroethane-d4	50	52.0	104	74	125
		Dibromofluoromethane	50	51.0	102	75	124
		Toluene-d8	50	49.0	98	86	113
		4-Bromofluorobenzene	50	49.0	98	77	121
VX0402WBS01	VX0402WBS01	1,2-Dichloroethane-d4	50	50.0	100	74	125
		Dibromofluoromethane	50	50.0	100	75	124
		Toluene-d8	50	50.0	100	86	113
		4-Bromofluorobenzene	50	49.0	98	77	121
VX0402WBSD0	VX0402WBSD01	1,2-Dichloroethane-d4	50	50.0	100	74	125
		Dibromofluoromethane	50	50.0	100	75	124
		Toluene-d8	50	50.0	100	86	113
		4-Bromofluorobenzene	50	51.0	102	77	121