

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q1623 SAS No.: Q1623 SDG NO.: Q1623
 Lab File ID: VX045534.D Date Analyzed: 04/02/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:02
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	75751	12.018				
UPPER LIMIT	151502	12.518				
LOWER LIMIT	37875.5	11.518				
EPA SAMPLE NO.						
Storage-Blank-SOIL-REF-6	52009	12.02				
Storage-Blank-WATER-REF-5	47363	12.02				
Storage-Blank-WATER-REF-4	47802	12.02				
Storage-Blank-SAMPLE-MANAGEMENT	53468	12.02				
VX0402WBL01	48101	12.02				
VX0402WBS01	68695	12.02				
VX0402WBSD01	66275	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.