

CASE NARRATIVE

Walsh Construction Company II, LLC
Project Name: Walsh CO-032 Sampling
Project # N/A
Chemtech Project # Q1626
Test Name: TCLP BNA

A. Number of Samples and Date of Receipt:

3 Solid samples were received on 03/21/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Ammonia, COD, Corrosivity, Cyanide, EPH_NF, Gasoline Range Organics, Herbicide, Hexavalent Chromium, Ignitability, Mercury, Metals ICP-TAL, METALS TAL+CN, METALS-TAL, Oil and Grease, Paint Filter, PCB, Pesticide-TCL, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SPLP BNA, SPLP Extraction, SPLP Herbicide, SPLP ICP Metals, SPLP Mercury, SPLP Pesticide, SPLP VOA, SPLP ZHE Ext, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP METALS, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, TPH GC, TS, TVS and VOC-TCLVOA-10. This data package contains results for TCLP BNA.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df. The analysis of TCLP BNA was based on method 8270E and extraction was done based on method 3510 and TCLP extraction method was 1311.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for CO-32-1 [Terphenyl-d14 - 127%]. As per method one surrogate is allowed to fail, Therefore no corrective action required.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS {Q1627-01MS} with File ID: BF142086.D recoveries met the requirements for all compounds except for 2,4,5-Trichlorophenol[118%], 2,4,6-Trichlorophenol[114%] and Hexachlorobenzene[124%] due to matrix interference .

The MSD {Q1627-01MSD} with File ID: BF142087.D recoveries met the acceptable requirements except for 2,4,5-Trichlorophenol[112%], 2,4,6-Trichlorophenol[118%] and Hexachlorobenzene[124%] due to matrix interference .

The RPD met criteria .

The Blank Spike for {PB167310BS} with File ID: BF142096.D met requirements for all samples except for 2,4-Dinitrotoluene[116%], Hexachlorobenzene[107%] and Hexachlorobutadiene[107%] . But associated samples have not positive hit for these compounds therefore no corrective action was taken.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

E. Additional Comments:

The Form 6 is not included in the data package because the Initial Calibration was performed using 8 points.

The soil samples results are based on a dry weight basis.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____