

Matrix Spike Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q1626
Project:	Walsh CO-032 Sampling	Sample ID:	Q1614-01
Client ID:	TR-04-032025MS	Percent Solids for Spike Sample:	94.6

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	75-125	1290		0.073	U	1360	40	95		03/24/2025

Matrix Spike Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q1626
Project:	Walsh CO-032 Sampling	Sample ID:	Q1614-01
Client ID:	TR-04-032025MS	Percent Solids for Spike Sample:	94.6

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	85-115	39.0		0.073	U	42.3	2	92		03/24/2025

Matrix Spike Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q1626
Project:	Walsh CO-032 Sampling	Sample ID:	Q1614-01
Client ID:	TR-04-032025MS	Percent Solids for Spike Sample:	94.6

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	75-125	33.0		0.073	U	42.3	2	78		03/24/2025

Matrix Spike Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q1626
Project:	Walsh CO-032 Sampling	Sample ID:	Q1626-01
Client ID:	CO-32-1MS	Percent Solids for Spike Sample:	94.8

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/Kg	75-125	1.70		0.043	U	2.1	1	81		03/25/2025

Matrix Spike Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q1626
Project:	Walsh CO-032 Sampling	Sample ID:	Q1626-01
Client ID:	CO-32-1MSD	Percent Solids for Spike Sample:	94.8

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Cyanide	mg/Kg	75-125	1.70		0.043	U	2.1	1	81		03/25/2025

Matrix Spike Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q1626
Project:	Walsh CO-032 Sampling	Sample ID:	Q1632-01
Client ID:	PIER-1-2MS	Percent Solids for Spike Sample:	95.4

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/Kg	75-125	36.6		20.9	J	105	1	15	*	03/25/2025
COD	mg/Kg	75-125	4570		2090		2570	1	96		03/25/2025
Ammonia as N	mg/Kg	75-125	53.2		2.20	U	51.4	1	104		03/25/2025

Matrix Spike Summary

Client:	Walsh Construction Company II, LLC	SDG No.:	Q1626
Project:	Walsh CO-032 Sampling	Sample ID:	Q1632-01
Client ID:	PIER-1-2MSD	Percent Solids for Spike Sample:	95.4

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Oil and Grease	mg/Kg	75-125	52.3		20.9	J	105	1	30	*	03/25/2025
COD	mg/Kg	75-125	4470		2090		2570	1	93		03/25/2025
Ammonia as N	mg/Kg	75-125	52.5		2.20	U	50.9	1	103		03/25/2025