

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135182

Review By	kareem	Review On	3/26/2025 10:50:02 AM
Supervise By	JANVI	Supervise On	3/26/2025 4:27:30 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84636,MP84637,MP84638,MP84639,MP84640,MP84846		
ICV Standard	MP84643		
CCV Standard	MP84646		
ICSA Standard	MP84644,MP84721		
CRI Standard	MP84846		
LCS Standard			
Chk Standard	MP84649,MP84650		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	03/25/25 09:38		Kareem	OK
2	S1	S1	CAL2	03/25/25 09:42		Kareem	OK
3	S2	S2	CAL3	03/25/25 09:46		Kareem	OK
4	S3	S3	CAL4	03/25/25 09:50		Kareem	OK
5	S4	S4	CAL5	03/25/25 09:54		Kareem	OK
6	S5	S5	CAL6	03/25/25 09:59		Kareem	OK
7	ICV01	ICV01	ICV	03/25/25 10:03		Kareem	OK
8	LLICV01	LLICV01	LLICV	03/25/25 10:14		Kareem	OK
9	ICB01	ICB01	ICB	03/25/25 10:18		Kareem	OK
10	CRI01	CRI01	CRDL	03/25/25 10:22		Kareem	OK
11	ICSA01	ICSA01	ICSA	03/25/25 10:27		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	03/25/25 10:31		Kareem	OK
13	CCV01	CCV01	CCV	03/25/25 10:35		Kareem	OK
14	CCB01	CCB01	CCB	03/25/25 10:40		Kareem	OK
15	Q1619-12	TP-4	SAM	03/25/25 10:44	Need to Confirm TCLP (Ba,Zn)	Kareem	OK
16	Q1619-14	TP-5	SAM	03/25/25 10:48	Need to Confirm TCLP (Ba,Zn)	Kareem	OK
17	Q1619-14DUP	TP-5DUP	DUP	03/25/25 10:53		Kareem	OK
18	Q1619-14L	TP-5L	SD	03/25/25 10:57		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135182

Review By	kareem	Review On	3/26/2025 10:50:02 AM
Supervise By	JANVI	Supervise On	3/26/2025 4:27:30 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84636,MP84637,MP84638,MP84639,MP84640,MP84846		
ICV Standard	MP84643		
CCV Standard	MP84646		
ICSA Standard	MP84644,MP84721		
CRI Standard	MP84846		
LCS Standard			
Chk Standard	MP84649,MP84650		

19	Q1619-14MS	TP-5MS	MS	03/25/25 11:01	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
20	Q1619-14MSD	TP-5MSD	MSD	03/25/25 11:06	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
21	Q1619-14A	TP-5A	PS	03/25/25 11:10	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
22	PB167260BL	PB167260BL	MB	03/25/25 11:14		Kareem	OK
23	PB167260BS	PB167260BS	LCS	03/25/25 11:18	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
24	Q1625-01	FRAC-TANK-A1291	SAM	03/25/25 11:22		Kareem	OK
25	CCV02	CCV02	CCV	03/25/25 11:27		Kareem	OK
26	CCB02	CCB02	CCB	03/25/25 11:31		Kareem	OK
27	Q1625-01DUP	FRAC-TANK-A1291D	DUP	03/25/25 11:35		Kareem	OK
28	Q1625-01L	FRAC-TANK-A1291L	SD	03/25/25 11:40		Kareem	OK
29	Q1625-01MS	FRAC-TANK-A1291M	MS	03/25/25 11:44	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
30	Q1625-01MSD	FRAC-TANK-A1291M	MSD	03/25/25 11:49	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
31	Q1625-01A	FRAC-TANK-A1291A	PS	03/25/25 11:53	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
32	PB167273BL	PB167273BL	MB	03/25/25 11:58		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135182

Review By	kareem	Review On	3/26/2025 10:50:02 AM
Supervise By	JANVI	Supervise On	3/26/2025 4:27:30 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84636,MP84637,MP84638,MP84639,MP84640,MP84846		
ICV Standard	MP84643		
CCV Standard	MP84646		
ICSA Standard	MP84644,MP84721		
CRI Standard	MP84846		
LCS Standard			
Chk Standard	MP84649,MP84650		

33	PB167273BS	PB167273BS	LCS	03/25/25 12:02	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
34	Q1610-02	SOIL-PILE	SAM	03/25/25 12:06		Kareem	OK
35	Q1619-01	CONCRETE-1	SAM	03/25/25 12:10		Kareem	OK
36	Q1619-03	CONCRETE-2	SAM	03/25/25 12:15		Kareem	OK
37	CCV03	CCV03	CCV	03/25/25 12:29		Kareem	OK
38	CCB03	CCB03	CCB	03/25/25 12:33		Kareem	OK
39	Q1619-05	TP-1	SAM	03/25/25 12:38	Cr high	Kareem	Dilution
40	Q1619-07	TP-2	SAM	03/25/25 12:42		Kareem	OK
41	Q1619-09	TP-3	SAM	03/25/25 12:46	Cr high	Kareem	Dilution
42	Q1619-11	TP-4	SAM	03/25/25 12:50	Cr high	Kareem	Dilution
43	Q1619-13	TP-5	SAM	03/25/25 12:54		Kareem	OK
44	PB167257BL	PB167257BL	MB	03/25/25 12:59		Kareem	OK
45	PB167257BS	PB167257BS	LCS	03/25/25 13:03		Kareem	OK
46	Q1624-01	OK-01-03212025	SAM	03/25/25 13:07		Kareem	OK
47	Q1624-01DUP	OK-01-03212025DUP	DUP	03/25/25 13:11		Kareem	OK
48	Q1624-01L	OK-01-03212025L	SD	03/25/25 13:15		Kareem	OK
49	CCV04	CCV04	CCV	03/25/25 13:33		Kareem	OK
50	CCB04	CCB04	CCB	03/25/25 13:42		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135182

Review By	kareem	Review On	3/26/2025 10:50:02 AM
Supervise By	JANVI	Supervise On	3/26/2025 4:27:30 PM

STD. NAME	STD REF.#
ICAL Standard	MP84636,MP84637,MP84638,MP84639,MP84640,MP84846
ICV Standard	MP84643
CCV Standard	MP84646
ICSA Standard	MP84644,MP84721
CRI Standard	MP84846
LCS Standard	
Chk Standard	MP84649,MP84650

51	Q1624-01MS	OK-01-03212025MS	MS	03/25/25 13:46	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
52	Q1624-01MSD	OK-01-03212025MSD	MSD	03/25/25 13:50	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
53	Q1624-01A	OK-01-03212025A	PS	03/25/25 13:54	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
54	Q1619-05DL	TP-1DL	SAM	03/25/25 13:58	10x for Cr	Kareem	Confirms
55	Q1626-01	CO-32-1	SAM	03/25/25 14:02		Kareem	OK
56	Q1630-01	VNJ-206	SAM	03/25/25 14:11		Kareem	OK
57	PB167267BL	PB167267BL	MB	03/25/25 14:15		Kareem	OK
58	PB167267BS	PB167267BS	LCS	03/25/25 14:20	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
59	Q1619-09DL	TP-3DL	SAM	03/25/25 14:24	25x for Cr	Kareem	Confirms
60	Q1619-11DL	TP-4DL	SAM	03/25/25 14:28	10x for Cr	Kareem	Confirms
61	CCV05	CCV05	CCV	03/25/25 14:38		Kareem	OK
62	CCB05	CCB05	CCB	03/25/25 14:48		Kareem	OK
63	PB167161BL	PB167161BL	MB	03/25/25 14:52		Kareem	OK
64	PB167133TB	PB167133TB	MB	03/25/25 15:00		Kareem	OK
65	PB167161BS	PB167161BS	LCS	03/25/25 15:12	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
66	Q1631-01	364	SAM	03/25/25 15:16		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135182

Review By	kareem	Review On	3/26/2025 10:50:02 AM
Supervise By	JANVI	Supervise On	3/26/2025 4:27:30 PM

STD. NAME	STD REF.#
ICAL Standard	MP84636,MP84637,MP84638,MP84639,MP84640,MP84846
ICV Standard	MP84643
CCV Standard	MP84646
ICSA Standard	MP84644,MP84721
CRI Standard	MP84846
LCS Standard	
Chk Standard	MP84649,MP84650

67	Q1631-04	359	SAM	03/25/25 15:21		Kareem	OK
68	Q1634-01	HR-01-032425	SAM	03/25/25 15:25		Kareem	OK
69	Q1634-03	HR-04-032425	SAM	03/25/25 15:29		Kareem	OK
70	Q1635-01	TP-2	SAM	03/25/25 15:33		Kareem	OK
71	Q1635-01DUP	TP-2DUP	DUP	03/25/25 15:37		Kareem	OK
72	Q1635-01L	TP-2L	SD	03/25/25 15:42		Kareem	OK
73	CCV06	CCV06	CCV	03/25/25 15:46		Kareem	OK
74	CCB06	CCB06	CCB	03/25/25 15:50		Kareem	OK
75	Q1635-01MS	TP-2MS	MS	03/25/25 15:54	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
76	Q1635-01MSD	TP-2MSD	MSD	03/25/25 15:58	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
77	Q1635-01A	TP-2A	PS	03/25/25 16:07	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
78	PB167296BL	PB167296BL	MB	03/25/25 16:11		Kareem	OK
79	PB167296BS	PB167296BS	LCS	03/25/25 16:15	0.1 ML OF M6004 AND M6013 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
80	Q1636-01	WC-1	SAM	03/25/25 16:19		Kareem	OK
81	Q1636-05	WC-2	SAM	03/25/25 16:23		Kareem	OK
82	Q1636-13	WC-4	SAM	03/25/25 16:27		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB135182

Review By	kareem	Review On	3/26/2025 10:50:02 AM
Supervise By	JANVI	Supervise On	3/26/2025 4:27:30 PM
STD. NAME	STD REF.#		
ICAL Standard	MP84636,MP84637,MP84638,MP84639,MP84640,MP84846		
ICV Standard	MP84643		
CCV Standard	MP84646		
ICSA Standard	MP84644,MP84721		
CRI Standard	MP84846		
LCS Standard			
Chk Standard	MP84649,MP84650		

83	Q1636-17	WC-5	SAM	03/25/25 16:31		Kareem	OK
84	Q1636-09	WC-3	SAM	03/25/25 16:35		Kareem	OK
85	CCV07	CCV07	CCV	03/25/25 16:40		Kareem	OK
86	CCB07	CCB07	CCB	03/25/25 16:44		Kareem	OK
87	LR1	LR1	HIGH STD	03/25/25 16:48		Kareem	OK
88	LR2	LR2	HIGH STD	03/25/25 16:53		Kareem	OK
89	CCV08	CCV08	CCV	03/25/25 17:01		Kareem	OK
90	CCB08	CCB08	CCB	03/25/25 17:05		Kareem	OK