

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Walsh Construction Company II, LLC **SDG No.:** Q1626
Contract: WALS01 **Lab Code:** CHEM **Case No.:** Q1626 **SAS No.:** Q1626
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV62	Mercury	4.01	4.0	100	90 - 110	CV	03/26/2025	11:23	LB135193

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Walsh Construction Company II, LLC **SDG No.:** Q1626
Contract: WALS01 **Lab Code:** CHEM **Case No.:** Q1626 **SAS No.:** Q1626
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV84	Mercury	5.12	5.0	102	90 - 110	CV	03/26/2025	11:32	LB135193
CCV85	Mercury	4.65	5.0	93	90 - 110	CV	03/26/2025	12:19	LB135193
CCV86	Mercury	4.65	5.0	93	90 - 110	CV	03/26/2025	12:50	LB135193
CCV87	Mercury	4.51	5.0	90	90 - 110	CV	03/26/2025	13:29	LB135193

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Walsh Construction Company II, LLC **SDG No.:** Q1626

Contract: WALS01 **Lab Code:** CHEM **Case No.:** Q1626 **SAS No.:** Q1626

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Aluminum	2460	2500	98	90 - 110	P	03/25/2025	10:03	LB135182
	Antimony	1000	1000	100	90 - 110	P	03/25/2025	10:03	LB135182
	Arsenic	989	1000	99	90 - 110	P	03/25/2025	10:03	LB135182
	Barium	493	520	95	90 - 110	P	03/25/2025	10:03	LB135182
	Beryllium	470	510	92	90 - 110	P	03/25/2025	10:03	LB135182
	Cadmium	494	510	97	90 - 110	P	03/25/2025	10:03	LB135182
	Calcium	9780	10000	98	90 - 110	P	03/25/2025	10:03	LB135182
	Chromium	518	520	100	90 - 110	P	03/25/2025	10:03	LB135182
	Cobalt	505	520	97	90 - 110	P	03/25/2025	10:03	LB135182
	Copper	526	510	103	90 - 110	P	03/25/2025	10:03	LB135182
	Iron	9990	10000	100	90 - 110	P	03/25/2025	10:03	LB135182
	Lead	975	1000	98	90 - 110	P	03/25/2025	10:03	LB135182
	Magnesium	5720	6000	95	90 - 110	P	03/25/2025	10:03	LB135182
	Manganese	498	520	96	90 - 110	P	03/25/2025	10:03	LB135182
	Nickel	508	530	96	90 - 110	P	03/25/2025	10:03	LB135182
	Potassium	9900	9900	100	90 - 110	P	03/25/2025	10:03	LB135182
	Selenium	1020	1000	102	90 - 110	P	03/25/2025	10:03	LB135182
	Silver	253	250	101	90 - 110	P	03/25/2025	10:03	LB135182
	Sodium	9650	10000	96	90 - 110	P	03/25/2025	10:03	LB135182
	Thallium	1040	1000	104	90 - 110	P	03/25/2025	10:03	LB135182
	Vanadium	483	500	96	90 - 110	P	03/25/2025	10:03	LB135182
	Zinc	1010	1000	101	90 - 110	P	03/25/2025	10:03	LB135182

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Walsh Construction Company II, LLC **SDG No.:** Q1626

Contract: WALS01 **Lab Code:** CHEM **Case No.:** Q1626 **SAS No.:** Q1626

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Aluminum	97.7	100	98	80 - 120	P	03/25/2025	10:14	LB135182
	Antimony	51.8	50.0	104	80 - 120	P	03/25/2025	10:14	LB135182
	Arsenic	18.5	20.0	93	80 - 120	P	03/25/2025	10:14	LB135182
	Barium	91.1	100	91	80 - 120	P	03/25/2025	10:14	LB135182
	Beryllium	5.83	6.0	97	80 - 120	P	03/25/2025	10:14	LB135182
	Cadmium	5.74	6.0	96	80 - 120	P	03/25/2025	10:14	LB135182
	Calcium	1980	2000	99	80 - 120	P	03/25/2025	10:14	LB135182
	Chromium	9.50	10.0	95	80 - 120	P	03/25/2025	10:14	LB135182
	Cobalt	29.2	30.0	97	80 - 120	P	03/25/2025	10:14	LB135182
	Copper	22.7	20.0	114	80 - 120	P	03/25/2025	10:14	LB135182
	Iron	95.4	100	95	80 - 120	P	03/25/2025	10:14	LB135182
	Lead	11.2	12.0	93	80 - 120	P	03/25/2025	10:14	LB135182
	Magnesium	2090	2000	104	80 - 120	P	03/25/2025	10:14	LB135182
	Manganese	18.8	20.0	94	80 - 120	P	03/25/2025	10:14	LB135182
	Nickel	39.0	40.0	98	80 - 120	P	03/25/2025	10:14	LB135182
	Potassium	1990	2000	100	80 - 120	P	03/25/2025	10:14	LB135182
	Selenium	21.7	20.0	108	80 - 120	P	03/25/2025	10:14	LB135182
	Silver	10.9	10.0	109	80 - 120	P	03/25/2025	10:14	LB135182
	Sodium	1870	2000	93	80 - 120	P	03/25/2025	10:14	LB135182
	Thallium	39.5	40.0	99	80 - 120	P	03/25/2025	10:14	LB135182
	Vanadium	39.3	40.0	98	80 - 120	P	03/25/2025	10:14	LB135182
	Zinc	43.0	40.0	108	80 - 120	P	03/25/2025	10:14	LB135182

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Walsh Construction Company II, LLC **SDG No.:** Q1626

Contract: WALS01 **Lab Code:** CHEM **Case No.:** Q1626 **SAS No.:** Q1626

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Aluminum	9850	10000	98	90 - 110	P	03/25/2025	10:35	LB135182
	Antimony	5180	5000	104	90 - 110	P	03/25/2025	10:35	LB135182
	Arsenic	5140	5000	103	90 - 110	P	03/25/2025	10:35	LB135182
	Barium	9690	10000	97	90 - 110	P	03/25/2025	10:35	LB135182
	Beryllium	237	250	95	90 - 110	P	03/25/2025	10:35	LB135182
	Cadmium	2470	2500	99	90 - 110	P	03/25/2025	10:35	LB135182
	Calcium	24000	25000	96	90 - 110	P	03/25/2025	10:35	LB135182
	Chromium	1010	1000	101	90 - 110	P	03/25/2025	10:35	LB135182
	Cobalt	2470	2500	99	90 - 110	P	03/25/2025	10:35	LB135182
	Copper	1280	1250	102	90 - 110	P	03/25/2025	10:35	LB135182
	Iron	5010	5000	100	90 - 110	P	03/25/2025	10:35	LB135182
	Lead	4980	5000	100	90 - 110	P	03/25/2025	10:35	LB135182
	Magnesium	24100	25000	96	90 - 110	P	03/25/2025	10:35	LB135182
	Manganese	2380	2500	95	90 - 110	P	03/25/2025	10:35	LB135182
	Nickel	2460	2500	98	90 - 110	P	03/25/2025	10:35	LB135182
	Potassium	25600	25000	102	90 - 110	P	03/25/2025	10:35	LB135182
	Selenium	5200	5000	104	90 - 110	P	03/25/2025	10:35	LB135182
	Silver	1260	1250	101	90 - 110	P	03/25/2025	10:35	LB135182
	Sodium	25100	25000	100	90 - 110	P	03/25/2025	10:35	LB135182
	Thallium	5100	5000	102	90 - 110	P	03/25/2025	10:35	LB135182
	Vanadium	2450	2500	98	90 - 110	P	03/25/2025	10:35	LB135182
	Zinc	2550	2500	102	90 - 110	P	03/25/2025	10:35	LB135182
CCV02	Aluminum	9680	10000	97	90 - 110	P	03/25/2025	11:27	LB135182
	Antimony	5000	5000	100	90 - 110	P	03/25/2025	11:27	LB135182
	Arsenic	4950	5000	99	90 - 110	P	03/25/2025	11:27	LB135182
	Barium	9380	10000	94	90 - 110	P	03/25/2025	11:27	LB135182
	Beryllium	234	250	94	90 - 110	P	03/25/2025	11:27	LB135182
	Cadmium	2390	2500	96	90 - 110	P	03/25/2025	11:27	LB135182
	Calcium	23200	25000	93	90 - 110	P	03/25/2025	11:27	LB135182
	Chromium	987	1000	99	90 - 110	P	03/25/2025	11:27	LB135182
	Cobalt	2400	2500	96	90 - 110	P	03/25/2025	11:27	LB135182
	Copper	1250	1250	100	90 - 110	P	03/25/2025	11:27	LB135182
	Iron	4820	5000	96	90 - 110	P	03/25/2025	11:27	LB135182
	Lead	4810	5000	96	90 - 110	P	03/25/2025	11:27	LB135182

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Walsh Construction Company II, LLC **SDG No.:** Q1626

Contract: WALS01 **Lab Code:** CHEM **Case No.:** Q1626 **SAS No.:** Q1626

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV02	Magnesium	23400	25000	94	90 - 110	P	03/25/2025	11:27	LB135182
	Manganese	2300	2500	92	90 - 110	P	03/25/2025	11:27	LB135182
	Nickel	2390	2500	96	90 - 110	P	03/25/2025	11:27	LB135182
	Potassium	24900	25000	100	90 - 110	P	03/25/2025	11:27	LB135182
	Selenium	4980	5000	100	90 - 110	P	03/25/2025	11:27	LB135182
	Silver	1230	1250	99	90 - 110	P	03/25/2025	11:27	LB135182
	Sodium	24300	25000	97	90 - 110	P	03/25/2025	11:27	LB135182
	Thallium	4940	5000	99	90 - 110	P	03/25/2025	11:27	LB135182
	Vanadium	2380	2500	95	90 - 110	P	03/25/2025	11:27	LB135182
	Zinc	2490	2500	99	90 - 110	P	03/25/2025	11:27	LB135182
CCV03	Aluminum	9630	10000	96	90 - 110	P	03/25/2025	12:29	LB135182
	Antimony	5090	5000	102	90 - 110	P	03/25/2025	12:29	LB135182
	Arsenic	5010	5000	100	90 - 110	P	03/25/2025	12:29	LB135182
	Barium	9540	10000	95	90 - 110	P	03/25/2025	12:29	LB135182
	Beryllium	226	250	90	90 - 110	P	03/25/2025	12:29	LB135182
	Cadmium	2390	2500	96	90 - 110	P	03/25/2025	12:29	LB135182
	Calcium	23100	25000	92	90 - 110	P	03/25/2025	12:29	LB135182
	Chromium	989	1000	99	90 - 110	P	03/25/2025	12:29	LB135182
	Cobalt	2410	2500	96	90 - 110	P	03/25/2025	12:29	LB135182
	Copper	1260	1250	101	90 - 110	P	03/25/2025	12:29	LB135182
	Iron	4970	5000	100	90 - 110	P	03/25/2025	12:29	LB135182
	Lead	4830	5000	97	90 - 110	P	03/25/2025	12:29	LB135182
	Magnesium	23000	25000	92	90 - 110	P	03/25/2025	12:29	LB135182
	Manganese	2300	2500	92	90 - 110	P	03/25/2025	12:29	LB135182
	Nickel	2400	2500	96	90 - 110	P	03/25/2025	12:29	LB135182
	Potassium	25800	25000	103	90 - 110	P	03/25/2025	12:29	LB135182
	Selenium	5070	5000	101	90 - 110	P	03/25/2025	12:29	LB135182
	Silver	1240	1250	99	90 - 110	P	03/25/2025	12:29	LB135182
	Sodium	25300	25000	101	90 - 110	P	03/25/2025	12:29	LB135182
	Thallium	4900	5000	98	90 - 110	P	03/25/2025	12:29	LB135182
	Vanadium	2380	2500	95	90 - 110	P	03/25/2025	12:29	LB135182
	Zinc	2510	2500	100	90 - 110	P	03/25/2025	12:29	LB135182
CCV04	Aluminum	9930	10000	99	90 - 110	P	03/25/2025	13:33	LB135182
	Antimony	5280	5000	106	90 - 110	P	03/25/2025	13:33	LB135182

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Walsh Construction Company II, LLC **SDG No.:** Q1626

Contract: WALS01 **Lab Code:** CHEM **Case No.:** Q1626 **SAS No.:** Q1626

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Arsenic	5180	5000	104	90 - 110	P	03/25/2025	13:33	LB135182
	Barium	9950	10000	100	90 - 110	P	03/25/2025	13:33	LB135182
	Beryllium	229	250	92	90 - 110	P	03/25/2025	13:33	LB135182
	Cadmium	2450	2500	98	90 - 110	P	03/25/2025	13:33	LB135182
	Calcium	23700	25000	95	90 - 110	P	03/25/2025	13:33	LB135182
	Chromium	1000	1000	100	90 - 110	P	03/25/2025	13:33	LB135182
	Cobalt	2470	2500	99	90 - 110	P	03/25/2025	13:33	LB135182
	Copper	1300	1250	104	90 - 110	P	03/25/2025	13:33	LB135182
	Iron	5140	5000	103	90 - 110	P	03/25/2025	13:33	LB135182
	Lead	4940	5000	99	90 - 110	P	03/25/2025	13:33	LB135182
	Magnesium	23500	25000	94	90 - 110	P	03/25/2025	13:33	LB135182
	Manganese	2360	2500	94	90 - 110	P	03/25/2025	13:33	LB135182
	Nickel	2460	2500	98	90 - 110	P	03/25/2025	13:33	LB135182
	Potassium	26700	25000	107	90 - 110	P	03/25/2025	13:33	LB135182
	Selenium	5260	5000	105	90 - 110	P	03/25/2025	13:33	LB135182
	Silver	1270	1250	102	90 - 110	P	03/25/2025	13:33	LB135182
	Sodium	26300	25000	105	90 - 110	P	03/25/2025	13:33	LB135182
	Thallium	5070	5000	102	90 - 110	P	03/25/2025	13:33	LB135182
	Vanadium	2450	2500	98	90 - 110	P	03/25/2025	13:33	LB135182
	Zinc	2590	2500	103	90 - 110	P	03/25/2025	13:33	LB135182
CCV05	Aluminum	9940	10000	99	90 - 110	P	03/25/2025	14:38	LB135182
	Antimony	5280	5000	106	90 - 110	P	03/25/2025	14:38	LB135182
	Arsenic	5180	5000	104	90 - 110	P	03/25/2025	14:38	LB135182
	Barium	9740	10000	97	90 - 110	P	03/25/2025	14:38	LB135182
	Beryllium	227	250	91	90 - 110	P	03/25/2025	14:38	LB135182
	Cadmium	2420	2500	97	90 - 110	P	03/25/2025	14:38	LB135182
	Calcium	23500	25000	94	90 - 110	P	03/25/2025	14:38	LB135182
	Chromium	1000	1000	100	90 - 110	P	03/25/2025	14:38	LB135182
	Cobalt	2440	2500	98	90 - 110	P	03/25/2025	14:38	LB135182
	Copper	1280	1250	103	90 - 110	P	03/25/2025	14:38	LB135182
	Iron	5080	5000	102	90 - 110	P	03/25/2025	14:38	LB135182
	Lead	4900	5000	98	90 - 110	P	03/25/2025	14:38	LB135182
	Magnesium	23500	25000	94	90 - 110	P	03/25/2025	14:38	LB135182
	Manganese	2330	2500	93	90 - 110	P	03/25/2025	14:38	LB135182

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Walsh Construction Company II, LLC **SDG No.:** Q1626

Contract: WALS01 **Lab Code:** CHEM **Case No.:** Q1626 **SAS No.:** Q1626

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Nickel	2430	2500	97	90 - 110	P	03/25/2025	14:38	LB135182
	Potassium	26700	25000	107	90 - 110	P	03/25/2025	14:38	LB135182
	Selenium	5260	5000	105	90 - 110	P	03/25/2025	14:38	LB135182
	Silver	1260	1250	101	90 - 110	P	03/25/2025	14:38	LB135182
	Sodium	25900	25000	104	90 - 110	P	03/25/2025	14:38	LB135182
	Thallium	4980	5000	100	90 - 110	P	03/25/2025	14:38	LB135182
	Vanadium	2440	2500	98	90 - 110	P	03/25/2025	14:38	LB135182
	Zinc	2560	2500	103	90 - 110	P	03/25/2025	14:38	LB135182
CCV06	Aluminum	9890	10000	99	90 - 110	P	03/25/2025	15:46	LB135182
	Antimony	5310	5000	106	90 - 110	P	03/25/2025	15:46	LB135182
	Arsenic	5190	5000	104	90 - 110	P	03/25/2025	15:46	LB135182
	Barium	9910	10000	99	90 - 110	P	03/25/2025	15:46	LB135182
	Beryllium	227	250	91	90 - 110	P	03/25/2025	15:46	LB135182
	Cadmium	2440	2500	98	90 - 110	P	03/25/2025	15:46	LB135182
	Calcium	23500	25000	94	90 - 110	P	03/25/2025	15:46	LB135182
	Chromium	1000	1000	100	90 - 110	P	03/25/2025	15:46	LB135182
	Cobalt	2460	2500	98	90 - 110	P	03/25/2025	15:46	LB135182
	Copper	1300	1250	104	90 - 110	P	03/25/2025	15:46	LB135182
	Iron	5100	5000	102	90 - 110	P	03/25/2025	15:46	LB135182
	Lead	4930	5000	98	90 - 110	P	03/25/2025	15:46	LB135182
	Magnesium	23300	25000	93	90 - 110	P	03/25/2025	15:46	LB135182
	Manganese	2340	2500	94	90 - 110	P	03/25/2025	15:46	LB135182
	Nickel	2450	2500	98	90 - 110	P	03/25/2025	15:46	LB135182
	Potassium	26800	25000	107	90 - 110	P	03/25/2025	15:46	LB135182
	Selenium	5270	5000	105	90 - 110	P	03/25/2025	15:46	LB135182
	Silver	1260	1250	101	90 - 110	P	03/25/2025	15:46	LB135182
	Sodium	26400	25000	106	90 - 110	P	03/25/2025	15:46	LB135182
	Thallium	4930	5000	99	90 - 110	P	03/25/2025	15:46	LB135182
	Vanadium	2440	2500	98	90 - 110	P	03/25/2025	15:46	LB135182
	Zinc	2570	2500	103	90 - 110	P	03/25/2025	15:46	LB135182
CCV07	Aluminum	9650	10000	96	90 - 110	P	03/25/2025	16:40	LB135182
	Antimony	5060	5000	101	90 - 110	P	03/25/2025	16:40	LB135182
	Arsenic	4970	5000	99	90 - 110	P	03/25/2025	16:40	LB135182
	Barium	9510	10000	95	90 - 110	P	03/25/2025	16:40	LB135182

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Walsh Construction Company II, LLC **SDG No.:** Q1626

Contract: WALS01 **Lab Code:** CHEM **Case No.:** Q1626 **SAS No.:** Q1626

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Beryllium	225	250	90	90 - 110	P	03/25/2025	16:40	LB135182
	Cadmium	2380	2500	95	90 - 110	P	03/25/2025	16:40	LB135182
	Calcium	22800	25000	91	90 - 110	P	03/25/2025	16:40	LB135182
	Chromium	972	1000	97	90 - 110	P	03/25/2025	16:40	LB135182
	Cobalt	2400	2500	96	90 - 110	P	03/25/2025	16:40	LB135182
	Copper	1250	1250	100	90 - 110	P	03/25/2025	16:40	LB135182
	Iron	4830	5000	97	90 - 110	P	03/25/2025	16:40	LB135182
	Lead	4770	5000	96	90 - 110	P	03/25/2025	16:40	LB135182
	Magnesium	22700	25000	91	90 - 110	P	03/25/2025	16:40	LB135182
	Manganese	2260	2500	90	90 - 110	P	03/25/2025	16:40	LB135182
	Nickel	2400	2500	96	90 - 110	P	03/25/2025	16:40	LB135182
	Potassium	25000	25000	100	90 - 110	P	03/25/2025	16:40	LB135182
	Selenium	5020	5000	100	90 - 110	P	03/25/2025	16:40	LB135182
	Silver	1220	1250	98	90 - 110	P	03/25/2025	16:40	LB135182
	Sodium	24400	25000	98	90 - 110	P	03/25/2025	16:40	LB135182
	Thallium	4770	5000	95	90 - 110	P	03/25/2025	16:40	LB135182
	Vanadium	2360	2500	94	90 - 110	P	03/25/2025	16:40	LB135182
	Zinc	2480	2500	99	90 - 110	P	03/25/2025	16:40	LB135182
CCV08	Aluminum	9700	10000	97	90 - 110	P	03/25/2025	17:01	LB135182
	Antimony	5070	5000	101	90 - 110	P	03/25/2025	17:01	LB135182
	Arsenic	5000	5000	100	90 - 110	P	03/25/2025	17:01	LB135182
	Barium	9410	10000	94	90 - 110	P	03/25/2025	17:01	LB135182
	Beryllium	234	250	94	90 - 110	P	03/25/2025	17:01	LB135182
	Cadmium	2410	2500	96	90 - 110	P	03/25/2025	17:01	LB135182
	Calcium	23200	25000	93	90 - 110	P	03/25/2025	17:01	LB135182
	Chromium	994	1000	99	90 - 110	P	03/25/2025	17:01	LB135182
	Cobalt	2430	2500	97	90 - 110	P	03/25/2025	17:01	LB135182
	Copper	1260	1250	101	90 - 110	P	03/25/2025	17:01	LB135182
	Iron	4880	5000	98	90 - 110	P	03/25/2025	17:01	LB135182
	Lead	4840	5000	97	90 - 110	P	03/25/2025	17:01	LB135182
	Magnesium	23300	25000	93	90 - 110	P	03/25/2025	17:01	LB135182
	Manganese	2300	2500	92	90 - 110	P	03/25/2025	17:01	LB135182
	Nickel	2430	2500	97	90 - 110	P	03/25/2025	17:01	LB135182
	Potassium	24900	25000	99	90 - 110	P	03/25/2025	17:01	LB135182

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Walsh Construction Company II, LLC **SDG No.:** Q1626
Contract: WALS01 **Lab Code:** CHEM **Case No.:** Q1626 **SAS No.:** Q1626
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV08	Selenium	5050	5000	101	90 - 110	P	03/25/2025	17:01	LB135182
	Silver	1240	1250	99	90 - 110	P	03/25/2025	17:01	LB135182
	Sodium	24000	25000	96	90 - 110	P	03/25/2025	17:01	LB135182
	Thallium	4810	5000	96	90 - 110	P	03/25/2025	17:01	LB135182
	Vanadium	2390	2500	96	90 - 110	P	03/25/2025	17:01	LB135182
	Zinc	2500	2500	100	90 - 110	P	03/25/2025	17:01	LB135182
