



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Walsh Construction Company II, LLC **SDG No.:** Q1626
Contract: WALS01 **Lab Code:** CHEM **Case No.:** Q1626 **SAS No.:** Q1626
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Aluminum	99.3	100	99	40 - 160	P	03/25/2025	10:22	LB135182
	Antimony	52.8	50.0	106	40 - 160	P	03/25/2025	10:22	LB135182
	Arsenic	21.2	20.0	106	40 - 160	P	03/25/2025	10:22	LB135182
	Barium	91.1	100	91	40 - 160	P	03/25/2025	10:22	LB135182
	Beryllium	5.76	6.0	96	40 - 160	P	03/25/2025	10:22	LB135182
	Cadmium	5.71	6.0	95	40 - 160	P	03/25/2025	10:22	LB135182
	Calcium	1960	2000	98	40 - 160	P	03/25/2025	10:22	LB135182
	Chromium	9.81	10.0	98	40 - 160	P	03/25/2025	10:22	LB135182
	Cobalt	28.9	30.0	96	40 - 160	P	03/25/2025	10:22	LB135182
	Copper	22.2	20.0	111	40 - 160	P	03/25/2025	10:22	LB135182
	Iron	98.2	100	98	40 - 160	P	03/25/2025	10:22	LB135182
	Lead	11.8	12.0	98	40 - 160	P	03/25/2025	10:22	LB135182
	Magnesium	2070	2000	104	40 - 160	P	03/25/2025	10:22	LB135182
	Manganese	18.7	20.0	93	40 - 160	P	03/25/2025	10:22	LB135182
	Nickel	39.0	40.0	98	40 - 160	P	03/25/2025	10:22	LB135182
	Potassium	2020	2000	101	40 - 160	P	03/25/2025	10:22	LB135182
	Selenium	21.9	20.0	110	40 - 160	P	03/25/2025	10:22	LB135182
	Silver	11.2	10.0	112	40 - 160	P	03/25/2025	10:22	LB135182
	Sodium	1880	2000	94	40 - 160	P	03/25/2025	10:22	LB135182
	Thallium	40.3	40.0	101	40 - 160	P	03/25/2025	10:22	LB135182
	Vanadium	39.5	40.0	99	40 - 160	P	03/25/2025	10:22	LB135182
	Zinc	42.3	40.0	106	40 - 160	P	03/25/2025	10:22	LB135182
CRA	Mercury	0.23	0.2	117	40 - 160	CV	03/26/2025	11:47	LB135193