



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Walsh Construction Company II, LLC</u>				SDG No.:	<u>Q1626</u>			
Contract:	<u>WALS01</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q1626</u>		SAS No.:	<u>Q1626</u>
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB62	Mercury	0.20	+/-0.20	U	0.20	CV	03/26/2025	11:29	LB135193

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q1626

Contract: WALS01

Lab Code: CHEM

Case No.: Q1626

SAS No.: Q1626

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB84	Mercury	0.20	+/-0.20	U	0.20	CV	03/26/2025	11:37	LB135193
CCB85	Mercury	0.20	+/-0.20	U	0.20	CV	03/26/2025	12:25	LB135193
CCB86	Mercury	0.20	+/-0.20	U	0.20	CV	03/26/2025	12:55	LB135193
CCB87	Mercury	0.20	+/-0.20	U	0.20	CV	03/26/2025	13:34	LB135193

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q1626

Contract: WALS01

Lab Code: CHEM

Case No.: Q1626

SAS No.: Q1626

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	03/25/2025	10:18	LB135182
	Antimony	50.0	+/-50.0	U	50.0	P	03/25/2025	10:18	LB135182
	Arsenic	20.0	+/-20.0	U	20.0	P	03/25/2025	10:18	LB135182
	Barium	100	+/-100	U	100	P	03/25/2025	10:18	LB135182
	Beryllium	6.00	+/-6.00	U	6.00	P	03/25/2025	10:18	LB135182
	Cadmium	6.00	+/-6.00	U	6.00	P	03/25/2025	10:18	LB135182
	Calcium	2000	+/-2000	U	2000	P	03/25/2025	10:18	LB135182
	Chromium	10.0	+/-10.0	U	10.0	P	03/25/2025	10:18	LB135182
	Cobalt	30.0	+/-30.0	U	30.0	P	03/25/2025	10:18	LB135182
	Copper	20.0	+/-20.0	U	20.0	P	03/25/2025	10:18	LB135182
	Iron	100	+/-100	U	100	P	03/25/2025	10:18	LB135182
	Lead	12.0	+/-12.0	U	12.0	P	03/25/2025	10:18	LB135182
	Magnesium	2000	+/-2000	U	2000	P	03/25/2025	10:18	LB135182
	Manganese	20.0	+/-20.0	U	20.0	P	03/25/2025	10:18	LB135182
	Nickel	40.0	+/-40.0	U	40.0	P	03/25/2025	10:18	LB135182
	Potassium	2000	+/-2000	U	2000	P	03/25/2025	10:18	LB135182
	Selenium	20.0	+/-20.0	U	20.0	P	03/25/2025	10:18	LB135182
	Silver	10.0	+/-10.0	U	10.0	P	03/25/2025	10:18	LB135182
	Sodium	2000	+/-2000	U	2000	P	03/25/2025	10:18	LB135182
	Thallium	40.0	+/-40.0	U	40.0	P	03/25/2025	10:18	LB135182
	Vanadium	40.0	+/-40.0	U	40.0	P	03/25/2025	10:18	LB135182
	Zinc	40.0	+/-40.0	U	40.0	P	03/25/2025	10:18	LB135182

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Walsh Construction Company II, LLC</u>		SDG No.: <u>Q1626</u>							
Contract: <u>WALS01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1626</u>	SAS No.: <u>Q1626</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	03/25/2025	10:40	LB135182
	Antimony	50.0	+/-50.0	U	50.0	P	03/25/2025	10:40	LB135182
	Arsenic	20.0	+/-20.0	U	20.0	P	03/25/2025	10:40	LB135182
	Barium	100	+/-100	U	100	P	03/25/2025	10:40	LB135182
	Beryllium	6.00	+/-6.00	U	6.00	P	03/25/2025	10:40	LB135182
	Cadmium	6.00	+/-6.00	U	6.00	P	03/25/2025	10:40	LB135182
	Calcium	2000	+/-2000	U	2000	P	03/25/2025	10:40	LB135182
	Chromium	10.0	+/-10.0	U	10.0	P	03/25/2025	10:40	LB135182
	Cobalt	30.0	+/-30.0	U	30.0	P	03/25/2025	10:40	LB135182
	Copper	20.0	+/-20.0	U	20.0	P	03/25/2025	10:40	LB135182
	Iron	100	+/-100	U	100	P	03/25/2025	10:40	LB135182
	Lead	12.0	+/-12.0	U	12.0	P	03/25/2025	10:40	LB135182
	Magnesium	2000	+/-2000	U	2000	P	03/25/2025	10:40	LB135182
	Manganese	20.0	+/-20.0	U	20.0	P	03/25/2025	10:40	LB135182
	Nickel	40.0	+/-40.0	U	40.0	P	03/25/2025	10:40	LB135182
	Potassium	2000	+/-2000	U	2000	P	03/25/2025	10:40	LB135182
	Selenium	20.0	+/-20.0	U	20.0	P	03/25/2025	10:40	LB135182
	Silver	10.0	+/-10.0	U	10.0	P	03/25/2025	10:40	LB135182
	Sodium	2000	+/-2000	U	2000	P	03/25/2025	10:40	LB135182
	Thallium	40.0	+/-40.0	U	40.0	P	03/25/2025	10:40	LB135182
	Vanadium	40.0	+/-40.0	U	40.0	P	03/25/2025	10:40	LB135182
	Zinc	40.0	+/-40.0	U	40.0	P	03/25/2025	10:40	LB135182
CCB02	Aluminum	100	+/-100	U	100	P	03/25/2025	11:31	LB135182
	Antimony	50.0	+/-50.0	U	50.0	P	03/25/2025	11:31	LB135182
	Arsenic	20.0	+/-20.0	U	20.0	P	03/25/2025	11:31	LB135182
	Barium	100	+/-100	U	100	P	03/25/2025	11:31	LB135182
	Beryllium	6.00	+/-6.00	U	6.00	P	03/25/2025	11:31	LB135182
	Cadmium	6.00	+/-6.00	U	6.00	P	03/25/2025	11:31	LB135182
	Calcium	2000	+/-2000	U	2000	P	03/25/2025	11:31	LB135182
	Chromium	10.0	+/-10.0	U	10.0	P	03/25/2025	11:31	LB135182
	Cobalt	30.0	+/-30.0	U	30.0	P	03/25/2025	11:31	LB135182
	Copper	20.0	+/-20.0	U	20.0	P	03/25/2025	11:31	LB135182
	Iron	100	+/-100	U	100	P	03/25/2025	11:31	LB135182
	Lead	12.0	+/-12.0	U	12.0	P	03/25/2025	11:31	LB135182
	Magnesium	2000	+/-2000	U	2000	P	03/25/2025	11:31	LB135182
	Manganese	20.0	+/-20.0	U	20.0	P	03/25/2025	11:31	LB135182
	Nickel	40.0	+/-40.0	U	40.0	P	03/25/2025	11:31	LB135182
	Potassium	2000	+/-2000	U	2000	P	03/25/2025	11:31	LB135182
	Selenium	20.0	+/-20.0	U	20.0	P	03/25/2025	11:31	LB135182

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q1626

Contract: WALS01

Lab Code: CHEM

Case No.: Q1626

SAS No.: Q1626

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	03/25/2025	11:31	LB135182
	Sodium	2000	+/-2000	U	2000	P	03/25/2025	11:31	LB135182
	Thallium	40.0	+/-40.0	U	40.0	P	03/25/2025	11:31	LB135182
	Vanadium	40.0	+/-40.0	U	40.0	P	03/25/2025	11:31	LB135182
	Zinc	40.0	+/-40.0	U	40.0	P	03/25/2025	11:31	LB135182
CCB03	Aluminum	100	+/-100	U	100	P	03/25/2025	12:33	LB135182
	Antimony	50.0	+/-50.0	U	50.0	P	03/25/2025	12:33	LB135182
	Arsenic	20.0	+/-20.0	U	20.0	P	03/25/2025	12:33	LB135182
	Barium	100	+/-100	U	100	P	03/25/2025	12:33	LB135182
	Beryllium	6.00	+/-6.00	U	6.00	P	03/25/2025	12:33	LB135182
	Cadmium	6.00	+/-6.00	U	6.00	P	03/25/2025	12:33	LB135182
	Calcium	2000	+/-2000	U	2000	P	03/25/2025	12:33	LB135182
	Chromium	10.0	+/-10.0	U	10.0	P	03/25/2025	12:33	LB135182
	Cobalt	30.0	+/-30.0	U	30.0	P	03/25/2025	12:33	LB135182
	Copper	20.0	+/-20.0	U	20.0	P	03/25/2025	12:33	LB135182
	Iron	100	+/-100	U	100	P	03/25/2025	12:33	LB135182
	Lead	12.0	+/-12.0	U	12.0	P	03/25/2025	12:33	LB135182
	Magnesium	2000	+/-2000	U	2000	P	03/25/2025	12:33	LB135182
	Manganese	20.0	+/-20.0	U	20.0	P	03/25/2025	12:33	LB135182
	Nickel	40.0	+/-40.0	U	40.0	P	03/25/2025	12:33	LB135182
	Potassium	2000	+/-2000	U	2000	P	03/25/2025	12:33	LB135182
	Selenium	20.0	+/-20.0	U	20.0	P	03/25/2025	12:33	LB135182
	Silver	10.0	+/-10.0	U	10.0	P	03/25/2025	12:33	LB135182
	Sodium	2000	+/-2000	U	2000	P	03/25/2025	12:33	LB135182
	Thallium	40.0	+/-40.0	U	40.0	P	03/25/2025	12:33	LB135182
	Vanadium	40.0	+/-40.0	U	40.0	P	03/25/2025	12:33	LB135182
	Zinc	40.0	+/-40.0	U	40.0	P	03/25/2025	12:33	LB135182
CCB04	Aluminum	100	+/-100	U	100	P	03/25/2025	13:42	LB135182
	Antimony	50.0	+/-50.0	U	50.0	P	03/25/2025	13:42	LB135182
	Arsenic	20.0	+/-20.0	U	20.0	P	03/25/2025	13:42	LB135182
	Barium	100	+/-100	U	100	P	03/25/2025	13:42	LB135182
	Beryllium	6.00	+/-6.00	U	6.00	P	03/25/2025	13:42	LB135182
	Cadmium	6.00	+/-6.00	U	6.00	P	03/25/2025	13:42	LB135182
	Calcium	2000	+/-2000	U	2000	P	03/25/2025	13:42	LB135182
	Chromium	10.0	+/-10.0	U	10.0	P	03/25/2025	13:42	LB135182
	Cobalt	30.0	+/-30.0	U	30.0	P	03/25/2025	13:42	LB135182
	Copper	20.0	+/-20.0	U	20.0	P	03/25/2025	13:42	LB135182
	Iron	100	+/-100	U	100	P	03/25/2025	13:42	LB135182
	Lead	12.0	+/-12.0	U	12.0	P	03/25/2025	13:42	LB135182

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Walsh Construction Company II, LLC</u>		SDG No.: <u>Q1626</u>							
Contract: <u>WALS01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1626</u>	SAS No.: <u>Q1626</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	2000	+/-2000	U	2000	P	03/25/2025	13:42	LB135182
	Manganese	20.0	+/-20.0	U	20.0	P	03/25/2025	13:42	LB135182
	Nickel	40.0	+/-40.0	U	40.0	P	03/25/2025	13:42	LB135182
	Potassium	2000	+/-2000	U	2000	P	03/25/2025	13:42	LB135182
	Selenium	20.0	+/-20.0	U	20.0	P	03/25/2025	13:42	LB135182
	Silver	10.0	+/-10.0	U	10.0	P	03/25/2025	13:42	LB135182
	Sodium	2000	+/-2000	U	2000	P	03/25/2025	13:42	LB135182
	Thallium	40.0	+/-40.0	U	40.0	P	03/25/2025	13:42	LB135182
	Vanadium	40.0	+/-40.0	U	40.0	P	03/25/2025	13:42	LB135182
	Zinc	40.0	+/-40.0	U	40.0	P	03/25/2025	13:42	LB135182
CCB05	Aluminum	100	+/-100	U	100	P	03/25/2025	14:48	LB135182
	Antimony	50.0	+/-50.0	U	50.0	P	03/25/2025	14:48	LB135182
	Arsenic	20.0	+/-20.0	U	20.0	P	03/25/2025	14:48	LB135182
	Barium	100	+/-100	U	100	P	03/25/2025	14:48	LB135182
	Beryllium	6.00	+/-6.00	U	6.00	P	03/25/2025	14:48	LB135182
	Cadmium	6.00	+/-6.00	U	6.00	P	03/25/2025	14:48	LB135182
	Calcium	2000	+/-2000	U	2000	P	03/25/2025	14:48	LB135182
	Chromium	10.0	+/-10.0	U	10.0	P	03/25/2025	14:48	LB135182
	Cobalt	30.0	+/-30.0	U	30.0	P	03/25/2025	14:48	LB135182
	Copper	20.0	+/-20.0	U	20.0	P	03/25/2025	14:48	LB135182
	Iron	100	+/-100	U	100	P	03/25/2025	14:48	LB135182
	Lead	12.0	+/-12.0	U	12.0	P	03/25/2025	14:48	LB135182
	Magnesium	2000	+/-2000	U	2000	P	03/25/2025	14:48	LB135182
	Manganese	20.0	+/-20.0	U	20.0	P	03/25/2025	14:48	LB135182
	Nickel	40.0	+/-40.0	U	40.0	P	03/25/2025	14:48	LB135182
	Potassium	2000	+/-2000	U	2000	P	03/25/2025	14:48	LB135182
	Selenium	20.0	+/-20.0	U	20.0	P	03/25/2025	14:48	LB135182
	Silver	10.0	+/-10.0	U	10.0	P	03/25/2025	14:48	LB135182
	Sodium	2000	+/-2000	U	2000	P	03/25/2025	14:48	LB135182
	Thallium	40.0	+/-40.0	U	40.0	P	03/25/2025	14:48	LB135182
	Vanadium	40.0	+/-40.0	U	40.0	P	03/25/2025	14:48	LB135182
	Zinc	40.0	+/-40.0	U	40.0	P	03/25/2025	14:48	LB135182
CCB06	Aluminum	100	+/-100	U	100	P	03/25/2025	15:50	LB135182
	Antimony	50.0	+/-50.0	U	50.0	P	03/25/2025	15:50	LB135182
	Arsenic	20.0	+/-20.0	U	20.0	P	03/25/2025	15:50	LB135182
	Barium	100	+/-100	U	100	P	03/25/2025	15:50	LB135182
	Beryllium	6.00	+/-6.00	U	6.00	P	03/25/2025	15:50	LB135182
	Cadmium	6.00	+/-6.00	U	6.00	P	03/25/2025	15:50	LB135182
	Calcium	2000	+/-2000	U	2000	P	03/25/2025	15:50	LB135182

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q1626

Contract: WALS01

Lab Code: CHEM

Case No.: Q1626

SAS No.: Q1626

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	10.0	+/-10.0	U	10.0	P	03/25/2025	15:50	LB135182
	Cobalt	30.0	+/-30.0	U	30.0	P	03/25/2025	15:50	LB135182
	Copper	20.0	+/-20.0	U	20.0	P	03/25/2025	15:50	LB135182
	Iron	100	+/-100	U	100	P	03/25/2025	15:50	LB135182
	Lead	12.0	+/-12.0	U	12.0	P	03/25/2025	15:50	LB135182
	Magnesium	2000	+/-2000	U	2000	P	03/25/2025	15:50	LB135182
	Manganese	20.0	+/-20.0	U	20.0	P	03/25/2025	15:50	LB135182
	Nickel	40.0	+/-40.0	U	40.0	P	03/25/2025	15:50	LB135182
	Potassium	2000	+/-2000	U	2000	P	03/25/2025	15:50	LB135182
	Selenium	20.0	+/-20.0	U	20.0	P	03/25/2025	15:50	LB135182
	Silver	10.0	+/-10.0	U	10.0	P	03/25/2025	15:50	LB135182
	Sodium	2000	+/-2000	U	2000	P	03/25/2025	15:50	LB135182
	Thallium	40.0	+/-40.0	U	40.0	P	03/25/2025	15:50	LB135182
	Vanadium	40.0	+/-40.0	U	40.0	P	03/25/2025	15:50	LB135182
	Zinc	40.0	+/-40.0	U	40.0	P	03/25/2025	15:50	LB135182
CCB07	Aluminum	100	+/-100	U	100	P	03/25/2025	16:44	LB135182
	Antimony	50.0	+/-50.0	U	50.0	P	03/25/2025	16:44	LB135182
	Arsenic	20.0	+/-20.0	U	20.0	P	03/25/2025	16:44	LB135182
	Barium	100	+/-100	U	100	P	03/25/2025	16:44	LB135182
	Beryllium	6.00	+/-6.00	U	6.00	P	03/25/2025	16:44	LB135182
	Cadmium	6.00	+/-6.00	U	6.00	P	03/25/2025	16:44	LB135182
	Calcium	2000	+/-2000	U	2000	P	03/25/2025	16:44	LB135182
	Chromium	10.0	+/-10.0	U	10.0	P	03/25/2025	16:44	LB135182
	Cobalt	30.0	+/-30.0	U	30.0	P	03/25/2025	16:44	LB135182
	Copper	20.0	+/-20.0	U	20.0	P	03/25/2025	16:44	LB135182
	Iron	100	+/-100	U	100	P	03/25/2025	16:44	LB135182
	Lead	12.0	+/-12.0	U	12.0	P	03/25/2025	16:44	LB135182
	Magnesium	2000	+/-2000	U	2000	P	03/25/2025	16:44	LB135182
	Manganese	20.0	+/-20.0	U	20.0	P	03/25/2025	16:44	LB135182
	Nickel	40.0	+/-40.0	U	40.0	P	03/25/2025	16:44	LB135182
	Potassium	2000	+/-2000	U	2000	P	03/25/2025	16:44	LB135182
	Selenium	20.0	+/-20.0	U	20.0	P	03/25/2025	16:44	LB135182
	Silver	10.0	+/-10.0	U	10.0	P	03/25/2025	16:44	LB135182
	Sodium	2000	+/-2000	U	2000	P	03/25/2025	16:44	LB135182
	Thallium	40.0	+/-40.0	U	40.0	P	03/25/2025	16:44	LB135182
	Vanadium	40.0	+/-40.0	U	40.0	P	03/25/2025	16:44	LB135182
	Zinc	40.0	+/-40.0	U	40.0	P	03/25/2025	16:44	LB135182
CCB08	Aluminum	100	+/-100	U	100	P	03/25/2025	17:05	LB135182
	Antimony	50.0	+/-50.0	U	50.0	P	03/25/2025	17:05	LB135182

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Walsh Construction Company II, LLC</u>		SDG No.: <u>Q1626</u>							
Contract: <u>WALS01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1626</u>	SAS No.: <u>Q1626</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	03/25/2025	17:05	LB135182
	Barium	100	+/-100	U	100	P	03/25/2025	17:05	LB135182
	Beryllium	6.00	+/-6.00	U	6.00	P	03/25/2025	17:05	LB135182
	Cadmium	6.00	+/-6.00	U	6.00	P	03/25/2025	17:05	LB135182
	Calcium	2000	+/-2000	U	2000	P	03/25/2025	17:05	LB135182
	Chromium	10.0	+/-10.0	U	10.0	P	03/25/2025	17:05	LB135182
	Cobalt	30.0	+/-30.0	U	30.0	P	03/25/2025	17:05	LB135182
	Copper	20.0	+/-20.0	U	20.0	P	03/25/2025	17:05	LB135182
	Iron	100	+/-100	U	100	P	03/25/2025	17:05	LB135182
	Lead	12.0	+/-12.0	U	12.0	P	03/25/2025	17:05	LB135182
	Magnesium	2000	+/-2000	U	2000	P	03/25/2025	17:05	LB135182
	Manganese	20.0	+/-20.0	U	20.0	P	03/25/2025	17:05	LB135182
	Nickel	40.0	+/-40.0	U	40.0	P	03/25/2025	17:05	LB135182
	Potassium	2000	+/-2000	U	2000	P	03/25/2025	17:05	LB135182
	Selenium	20.0	+/-20.0	U	20.0	P	03/25/2025	17:05	LB135182
	Silver	10.0	+/-10.0	U	10.0	P	03/25/2025	17:05	LB135182
	Sodium	2000	+/-2000	U	2000	P	03/25/2025	17:05	LB135182
	Thallium	40.0	+/-40.0	U	40.0	P	03/25/2025	17:05	LB135182
	Vanadium	40.0	+/-40.0	U	40.0	P	03/25/2025	17:05	LB135182
	Zinc	40.0	+/-40.0	U	40.0	P	03/25/2025	17:05	LB135182



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Walsh Construction Company II, LLC</u>				SDG No.:	<u>Q1626</u>			
Contract:	<u>WALS01</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1626</u>	SAS No.:	<u>Q1626</u>		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number