

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC

SDG No.: Q1626

Contract: WALS01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Aluminum	235000	255000	92	216000	294000	03/25/2025	10:27	LB135182
	Antimony	-3.08			-50	50	03/25/2025	10:27	LB135182
	Arsenic	4.77			-20	20	03/25/2025	10:27	LB135182
	Barium	-0.56	6.0	9	-94	106	03/25/2025	10:27	LB135182
	Beryllium	1.21			-6	6	03/25/2025	10:27	LB135182
	Cadmium	-3.57	1.0	357	-5	7	03/25/2025	10:27	LB135182
	Calcium	221000	245000	90	208000	282000	03/25/2025	10:27	LB135182
	Chromium	55.0	52.0	106	42	62	03/25/2025	10:27	LB135182
	Cobalt	1.66			-30	30	03/25/2025	10:27	LB135182
	Copper	9.59	2.0	480	-18	22	03/25/2025	10:27	LB135182
	Iron	97800	101000	97	85600	116500	03/25/2025	10:27	LB135182
	Lead	4.19			-12	12	03/25/2025	10:27	LB135182
	Magnesium	236000	255000	92	216000	294000	03/25/2025	10:27	LB135182
	Manganese	1.54	7.0	22	-13	27	03/25/2025	10:27	LB135182
	Nickel	2.08	2.0	104	-38	42	03/25/2025	10:27	LB135182
	Potassium	46.7			0	0	03/25/2025	10:27	LB135182
	Selenium	-9.76			-20	20	03/25/2025	10:27	LB135182
	Silver	-0.0040			-10	10	03/25/2025	10:27	LB135182
	Sodium	53.4			0	0	03/25/2025	10:27	LB135182
	Thallium	5.68			-40	40	03/25/2025	10:27	LB135182
	Vanadium	4.19			-40	40	03/25/2025	10:27	LB135182
	Zinc	3.53			-40	40	03/25/2025	10:27	LB135182
ICSAB01	Aluminum	242000	247000	98	209000	285000	03/25/2025	10:31	LB135182
	Antimony	641	618	104	525	711	03/25/2025	10:31	LB135182
	Arsenic	112	104	108	88.4	120	03/25/2025	10:31	LB135182
	Barium	469	537	87	437	637	03/25/2025	10:31	LB135182
	Beryllium	480	495	97	420	570	03/25/2025	10:31	LB135182
	Cadmium	998	972	103	826	1120	03/25/2025	10:31	LB135182
	Calcium	224000	235000	95	199000	271000	03/25/2025	10:31	LB135182
	Chromium	566	542	104	460	624	03/25/2025	10:31	LB135182
	Cobalt	507	476	106	404	548	03/25/2025	10:31	LB135182
	Copper	516	511	101	434	588	03/25/2025	10:31	LB135182
	Iron	97700	99300	98	84400	114500	03/25/2025	10:31	LB135182
	Lead	52.0	49.0	106	37	61	03/25/2025	10:31	LB135182
	Magnesium	244000	248000	98	210000	286000	03/25/2025	10:31	LB135182
	Manganese	466	507	92	430	584	03/25/2025	10:31	LB135182
	Nickel	1000	954	105	810	1100	03/25/2025	10:31	LB135182
	Potassium	-12.6			0	0	03/25/2025	10:31	LB135182
	Selenium	37.0	46.0	80	26	66	03/25/2025	10:31	LB135182
	Silver	203	201	101	170	232	03/25/2025	10:31	LB135182
	Sodium	20.8			0	0	03/25/2025	10:31	LB135182
	Thallium	106	108	98	68	148	03/25/2025	10:31	LB135182

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC

SDG No.: Q1626

Contract: WALS01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSAB01	Vanadium	469	491	96	417	565	03/25/2025	10:31	LB135182
	Zinc	1070	952	112	809	1095	03/25/2025	10:31	LB135182