

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q1626

Contract: WALS01

Lab Code: CHEM

Case No.: Q1626

SAS No.: Q1626

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167267BS							
Aluminum	mg/Kg	95.7	89.0		93	80 - 120	P
Antimony	mg/Kg	38.3	39.0		102	80 - 120	P
Arsenic	mg/Kg	38.3	36.8		96	80 - 120	P
Barium	mg/Kg	9.6	8.00		83	80 - 120	P
Beryllium	mg/Kg	9.6	8.14		85	80 - 120	P
Cadmium	mg/Kg	9.6	8.61		90	80 - 120	P
Calcium	mg/Kg	47.8	43.8	J	92	80 - 120	P
Chromium	mg/Kg	19.1	18.4		96	80 - 120	P
Cobalt	mg/Kg	9.6	8.81		92	80 - 120	P
Copper	mg/Kg	14.4	14.6		101	80 - 120	P
Iron	mg/Kg	140	137		98	80 - 120	P
Lead	mg/Kg	47.8	43.3		91	80 - 120	P
Magnesium	mg/Kg	95.7	82.9	J	87	80 - 120	P
Manganese	mg/Kg	9.6	8.44		88	80 - 120	P
Nickel	mg/Kg	23.9	22.1		92	80 - 120	P
Potassium	mg/Kg	480	463		96	80 - 120	P
Selenium	mg/Kg	95.7	95.2		100	80 - 120	P
Silver	mg/Kg	3.6	3.50		97	80 - 120	P
Sodium	mg/Kg	140	131		94	80 - 120	P
Thallium	mg/Kg	95.7	90.7		95	80 - 120	P
Vanadium	mg/Kg	14.4	13.0		90	80 - 120	P
Zinc	mg/Kg	9.6	9.39		98	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Walsh Construction Company II, LLC

SDG No.: Q1626

Contract: WALS01

Lab Code: CHEM

Case No.: Q1626

SAS No.: Q1626

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167317BS Mercury	mg/Kg	0.25	0.25		100	80 - 120	CV