



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Walsh Construction Company II, LLC **SDG No.:** Q1626
Contract: WALS01 **Lab Code:** CHEM **Case No.:** Q1626 **SAS No.:** Q1626
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.17	0.2	85	40 - 160	CV	03/26/2025	08:48	LB135183
CRI01	Arsenic	19.5	20.0	97	40 - 160	P	03/26/2025	11:22	LB135196
	Barium	88.3	100	88	40 - 160	P	03/26/2025	11:22	LB135196
	Cadmium	5.73	6.0	96	40 - 160	P	03/26/2025	11:22	LB135196
	Chromium	10.1	10.0	101	40 - 160	P	03/26/2025	11:22	LB135196
	Lead	10.9	12.0	91	40 - 160	P	03/26/2025	11:22	LB135196
	Selenium	19.6	20.0	98	40 - 160	P	03/26/2025	11:22	LB135196
	Silver	11.0	10.0	110	40 - 160	P	03/26/2025	11:22	LB135196