

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Walsh Construction Company II, LLC **SDG No.:** Q1626
Contract: WALS01 **Lab Code:** CHEM **Case No.:** Q1626 **SAS No.:** Q1626
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	4.64			-20	20	03/26/2025	11:27	LB135196
	Barium	0.78	6.0	13	-94	106	03/26/2025	11:27	LB135196
	Cadmium	-2.74	1.0	274	-5	7	03/26/2025	11:27	LB135196
	Chromium	57.2	52.0	110	42	62	03/26/2025	11:27	LB135196
	Lead	11.8			-12	12	03/26/2025	11:27	LB135196
	Selenium	-16.2			-20	20	03/26/2025	11:27	LB135196
	Silver	-0.98			-10	10	03/26/2025	11:27	LB135196
ICSAB01	Arsenic	110	104	106	88.4	120	03/26/2025	11:31	LB135196
	Barium	483	537	90	437	637	03/26/2025	11:31	LB135196
	Cadmium	1020	972	105	826	1120	03/26/2025	11:31	LB135196
	Chromium	579	542	107	460	624	03/26/2025	11:31	LB135196
	Lead	58.7	49.0	120	37	61	03/26/2025	11:31	LB135196
	Selenium	35.0	46.0	76	26	66	03/26/2025	11:31	LB135196
	Silver	205	201	102	170	232	03/26/2025	11:31	LB135196