

Metals
-9 -
ICP SERIAL DILUTIONS

SAMPLE NO.

WC-5L

Lab Name: Chemtech Consulting Group

Contract: WAL501

Lab Code: CHEM **Lb No.:** lb135196 **Lab Sample ID :** Q1636-20L **SDG No.:** Q1626

Matrix (soil/water): Water **Level (low/med):** LOW

Concentration Units: ug/L

Analyte	Initial Sample Result (I)		Serial Dilution Result (S)		% Difference	Q	M
		C		C			
Arsenic	100	U	500	U			P
Barium	1850		1750	J	5		P
Cadmium	30.0	U	150	U			P
Chromium	50.0	U	250	U			P
Lead	60.0	U	300	U			P
Mercury	2.00	U	10.0	U			CV
Selenium	100	U	500	U			P
Silver	50.0	U	250	U			P