

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS032725

Review By	Abdul	Review On	3/27/2025 8:26:06 AM
Supervise By	mohammad	Supervise On	3/28/2025 1:18:23 AM
SubDirectory	PS032725	HP Acquire Method	HP Processing Method ps032527 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24066 PP24069,PP24070

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS029542.D	26 Mar 2025 11:48		AR\AJ	Ok
2	I.BLK	I.BLK	PS029543.D	26 Mar 2025 12:12		AR\AJ	Ok,M
3	HSTDICC200	HSTDICC200	PS029544.D	26 Mar 2025 12:36	Not in Use	AR\AJ	Not Ok
4	HSTDICC500	HSTDICC500	PS029545.D	26 Mar 2025 13:04	Method Fail	AR\AJ	Not Ok
5	HSTDICC750	HSTDICC750	PS029546.D	26 Mar 2025 13:28	Method Fail	AR\AJ	Not Ok
6	HSTDICC1000	HSTDICC1000	PS029547.D	26 Mar 2025 13:52	Method Fail	AR\AJ	Not Ok
7	HSTDICC1500	HSTDICC1500	PS029548.D	26 Mar 2025 14:16	Method Fail	AR\AJ	Not Ok
8	HSTDICC200	HSTDICC200	PS029549.D	26 Mar 2025 15:28	Method Fail, F flag in comp#15	AR\AJ	Not Ok
9	HSTDICV750	ICVPS032725	PS029550.D	26 Mar 2025 16:06	Method Fail	AR\AJ	Not Ok
10	I.BLK	I.BLK	PS029551.D	26 Mar 2025 16:30		AR\AJ	Ok
11	HSTDCCC750	HSTDCCC750	PS029552.D	26 Mar 2025 16:54	Method fail	AR\AJ	Not Ok
12	Q1627-01	GRID-LINE-1.2-NORTH	PS029553.D	26 Mar 2025 17:18	Method fail	AR\AJ	ReRun
13	Q1626-01	CO-32-1	PS029554.D	26 Mar 2025 17:42	Method fail	AR\AJ	ReRun
14	Q1626-03	CO-32-1	PS029555.D	26 Mar 2025 18:06	Method fail	AR\AJ	ReRun

M : Manual Integration