

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 03/26/2025 03/26/2025

Client Sample No.: CCAL01 **Date Analyzed:** 03/26/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS029552.D **Time Analyzed:** 16:54

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.381	9.282	9.482	827.740	712.500	16.2
2,4,5-TP(Silvex)	9.088	8.988	9.188	752.360	712.500	5.6
2,4-D	8.230	8.131	8.331	803.930	705.000	14.0
2,4-DB	9.952	9.853	10.053	913.600	712.500	28.2
2,4-DCAA	7.121	7.021	7.221	783.910	750.000	4.5
DICAMBA	7.300	7.200	7.400	725.340	705.000	2.9
DICHLORPROP	7.997	7.896	8.096	727.380	705.000	3.2
Dinoseb	11.132	11.032	11.232	752.480	705.000	6.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 03/26/2025 03/26/2025

Client Sample No.: CCAL01 **Date Analyzed:** 03/26/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS029552.D **Time Analyzed:** 16:54

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.050	9.950	10.150	751.220	712.500	5.4
2,4,5-TP(Silvex)	9.639	9.539	9.739	749.540	712.500	5.2
2,4-D	8.755	8.654	8.854	726.720	705.000	3.1
2,4-DB	10.609	10.509	10.709	742.130	712.500	4.2
2,4-DCAA	7.542	7.442	7.642	774.480	750.000	3.3
DICAMBA	7.731	7.631	7.831	741.410	705.000	5.2
DICHLORPROP	8.434	8.333	8.533	726.770	705.000	3.1
Dinoseb	10.981	10.881	11.081	734.420	705.000	4.2