

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 03/27/2025 03/27/2025

Client Sample No.: CCAL02 **Date Analyzed:** 03/27/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS029588.D **Time Analyzed:** 22:51

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.371	9.273	9.473	761.260	712.500	6.8
2,4,5-TP(Silvex)	9.080	8.982	9.182	718.400	712.500	0.8
2,4-D	8.220	8.122	8.322	741.930	705.000	5.2
2,4-DB	9.939	9.843	10.043	808.180	712.500	13.4
2,4-DCAA	7.114	7.015	7.215	743.510	750.000	-0.9
DICAMBA	7.293	7.194	7.394	691.940	705.000	-1.9
DICHLORPROP	7.990	7.891	8.091	693.070	705.000	-1.7
Dinoseb	11.124	11.026	11.226	706.110	705.000	0.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 03/27/2025 03/27/2025

Client Sample No.: CCAL02 **Date Analyzed:** 03/27/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS029588.D **Time Analyzed:** 22:51

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.040	9.943	10.143	731.070	712.500	2.6
2,4,5-TP(Silvex)	9.630	9.534	9.734	728.840	712.500	2.3
2,4-D	8.745	8.648	8.848	714.600	705.000	1.4
2,4-DB	10.598	10.502	10.702	657.450	712.500	-7.7
2,4-DCAA	7.534	7.437	7.637	753.070	750.000	0.4
DICAMBA	7.724	7.626	7.826	723.310	705.000	2.6
DICHLORPROP	8.425	8.328	8.528	690.810	705.000	-2.0
Dinoseb	10.971	10.874	11.074	707.320	705.000	0.3