

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 03/27/2025 03/27/2025

Client Sample No.: CCAL04 **Date Analyzed:** 03/28/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS029612.D **Time Analyzed:** 09:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.372	9.273	9.473	785.690	712.500	10.3
2,4,5-TP(Silvex)	9.081	8.982	9.182	742.110	712.500	4.2
2,4-D	8.220	8.122	8.322	764.850	705.000	8.5
2,4-DB	9.940	9.843	10.043	855.010	712.500	20.0
2,4-DCAA	7.114	7.015	7.215	761.470	750.000	1.5
DICAMBA	7.293	7.194	7.394	713.270	705.000	1.2
DICHLORPROP	7.990	7.891	8.091	718.940	705.000	2.0
Dinoseb	11.126	11.026	11.226	733.150	705.000	4.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 03/27/2025 03/27/2025

Client Sample No.: CCAL04 **Date Analyzed:** 03/28/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS029612.D **Time Analyzed:** 09:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	10.038	9.943	10.143	749.100	712.500	5.1
2,4,5-TP(Silvex)	9.628	9.534	9.734	747.960	712.500	5.0
2,4-D	8.744	8.648	8.848	727.670	705.000	3.2
2,4-DB	10.596	10.502	10.702	663.410	712.500	-6.9
2,4-DCAA	7.533	7.437	7.637	767.240	750.000	2.3
DICAMBA	7.722	7.626	7.826	741.650	705.000	5.2
DICHLORPROP	8.424	8.328	8.528	713.010	705.000	1.1
Dinoseb	10.969	10.874	11.074	737.540	705.000	4.6