

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 04/02/2025 04/02/2025

Client Sample No.: CCAL06 **Date Analyzed:** 04/03/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS029683.D **Time Analyzed:** 17:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.167	9.070	9.270	758.130	712.500	6.4
2,4,5-TP(Silvex)	8.885	8.788	8.988	749.150	712.500	5.1
2,4-D	8.039	7.942	8.142	743.970	705.000	5.5
2,4-DB	9.726	9.629	9.829	794.180	712.500	11.5
2,4-DCAA	6.959	6.863	7.063	774.190	750.000	3.2
DICAMBA	7.136	7.040	7.240	756.430	705.000	7.3
DICHLORPROP	7.818	7.722	7.922	716.230	705.000	1.6
Dinoseb	10.891	10.795	10.995	739.250	705.000	4.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 04/02/2025 04/02/2025

Client Sample No.: CCAL06 **Date Analyzed:** 04/03/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS029683.D **Time Analyzed:** 17:59

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.952	9.857	10.057	750.240	712.500	5.3
2,4,5-TP(Silvex)	9.547	9.451	9.651	739.400	712.500	3.8
2,4-D	8.667	8.571	8.771	698.790	705.000	-0.9
2,4-DB	10.509	10.414	10.614	734.590	712.500	3.1
2,4-DCAA	7.472	7.375	7.575	761.460	750.000	1.5
DICAMBA	7.658	7.562	7.762	729.850	705.000	3.5
DICHLORPROP	8.353	8.256	8.456	695.620	705.000	-1.3
Dinoseb	10.881	10.787	10.987	725.750	705.000	2.9