

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS032725

Review By	Abdul	Review On	3/27/2025 8:26:06 AM
Supervise By	mohammad	Supervise On	3/28/2025 1:18:23 AM
SubDirectory	PS032725	HP Acquire Method	HP Processing Method ps032527 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068		
CCC Internal Standard/PEM	PP24066		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS029542.D	26 Mar 2025 11:48	AR\AJ	Ok
2	I.BLK	PS029543.D	26 Mar 2025 12:12	AR\AJ	Ok,M
3	HSTDICC200	PS029544.D	26 Mar 2025 12:36	AR\AJ	Not Ok
4	HSTDICC500	PS029545.D	26 Mar 2025 13:04	AR\AJ	Not Ok
5	HSTDICC750	PS029546.D	26 Mar 2025 13:28	AR\AJ	Not Ok
6	HSTDICC1000	PS029547.D	26 Mar 2025 13:52	AR\AJ	Not Ok
7	HSTDICC1500	PS029548.D	26 Mar 2025 14:16	AR\AJ	Not Ok
8	HSTDICC200	PS029549.D	26 Mar 2025 15:28	AR\AJ	Not Ok
9	HSTDICV750	PS029550.D	26 Mar 2025 16:06	AR\AJ	Not Ok
10	I.BLK	PS029551.D	26 Mar 2025 16:30	AR\AJ	Ok
11	HSTDCCC750	PS029552.D	26 Mar 2025 16:54	AR\AJ	Not Ok
12	Q1627-01	PS029553.D	26 Mar 2025 17:18	AR\AJ	ReRun
13	Q1626-01	PS029554.D	26 Mar 2025 17:42	AR\AJ	ReRun
14	Q1626-03	PS029555.D	26 Mar 2025 18:06	AR\AJ	ReRun

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS032825

Review By	Abdul	Review On	3/28/2025 8:31:32 AM
Supervise By	mohammad	Supervise On	4/1/2025 1:29:30 AM
SubDirectory	PS032825	HP Acquire Method	HP Processing Method ps032825 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068		
CCC Internal Standard/PEM	PP24066		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS029556.D	27 Mar 2025 09:25	AR\AJ	Ok
2	I.BLK	PS029557.D	27 Mar 2025 09:49	AR\AJ	Ok
3	HSTDICC200	PS029558.D	27 Mar 2025 10:13	AR\AJ	Ok
4	HSTDICC500	PS029559.D	27 Mar 2025 10:37	AR\AJ	Ok
5	HSTDICC750	PS029560.D	27 Mar 2025 11:01	AR\AJ	Ok
6	HSTDICC1000	PS029561.D	27 Mar 2025 11:25	AR\AJ	Ok
7	HSTDICC1500	PS029562.D	27 Mar 2025 11:49	AR\AJ	Ok
8	HSTDICV750	PS029563.D	27 Mar 2025 12:25	AR\AJ	Ok
9	I.BLK	PS029564.D	27 Mar 2025 12:49	AR\AJ	Ok
10	HSTDCCC750	PS029565.D	27 Mar 2025 13:13	AR\AJ	Ok
11	Q1609-03	PS029566.D	27 Mar 2025 13:37	AR\AJ	Ok,M
12	Q1625-01	PS029567.D	27 Mar 2025 14:01	AR\AJ	Ok
13	Q1632-02	PS029568.D	27 Mar 2025 14:25	AR\AJ	Ok
14	Q1630-01	PS029569.D	27 Mar 2025 14:49	AR\AJ	Ok,M
15	Q1636-01	PS029570.D	27 Mar 2025 15:13	AR\AJ	Ok,M
16	Q1636-05	PS029571.D	27 Mar 2025 15:38	AR\AJ	Ok,M
17	Q1636-09	PS029572.D	27 Mar 2025 16:02	AR\AJ	Ok
18	Q1636-13	PS029573.D	27 Mar 2025 16:26	AR\AJ	Ok
19	Q1636-17	PS029574.D	27 Mar 2025 16:50	AR\AJ	Ok,M
20	I.BLK	PS029575.D	27 Mar 2025 17:14	AR\AJ	Ok,M
21	HSTDCCC750	PS029576.D	27 Mar 2025 18:02	AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS032825

Review By	Abdul	Review On	3/28/2025 8:31:32 AM
Supervise By	mohammad	Supervise On	4/1/2025 1:29:30 AM
SubDirectory	PS032825	HP Acquire Method	HP Processing Method ps032825 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24066 PP24069,PP24070		

22	Q1609-03MS	PS029577.D	27 Mar 2025 18:26	AR\AJ	Ok,M
23	Q1609-03MSD	PS029578.D	27 Mar 2025 18:50	AR\AJ	Ok,M
24	Q1636-05MS	PS029579.D	27 Mar 2025 19:14	AR\AJ	Ok,M
25	Q1636-05MSD	PS029580.D	27 Mar 2025 19:38	AR\AJ	Ok,M
26	Q1648-01	PS029581.D	27 Mar 2025 20:02	AR\AJ	Ok
27	Q1648-03	PS029582.D	27 Mar 2025 20:26	AR\AJ	Ok,M
28	Q1648-05	PS029583.D	27 Mar 2025 20:50	AR\AJ	Ok,M
29	Q1648-07	PS029584.D	27 Mar 2025 21:14	AR\AJ	Ok,M
30	Q1648-09	PS029585.D	27 Mar 2025 21:38	AR\AJ	Ok,M
31	Q1648-11	PS029586.D	27 Mar 2025 22:02	AR\AJ	Ok,M
32	I.BLK	PS029587.D	27 Mar 2025 22:26	AR\AJ	Ok,M
33	HSTDCCC750	PS029588.D	27 Mar 2025 22:51	AR\AJ	Ok,M
34	Q1648-13	PS029589.D	27 Mar 2025 23:39	AR\AJ	Ok,M
35	Q1648-13MS	PS029590.D	28 Mar 2025 00:03	AR\AJ	Ok,M
36	Q1648-13MSD	PS029591.D	28 Mar 2025 00:27	AR\AJ	Ok,M
37	Q1627-01	PS029592.D	28 Mar 2025 00:51	AR\AJ	Ok,M
38	Q1626-01	PS029593.D	28 Mar 2025 01:15	AR\AJ	Ok,M
39	Q1626-01MS	PS029594.D	28 Mar 2025 01:39	AR\AJ	Ok,M
40	Q1626-01MSD	PS029595.D	28 Mar 2025 02:03	AR\AJ	Ok,M
41	Q1626-03	PS029596.D	28 Mar 2025 02:27	AR\AJ	Ok,M
42	PB167312BL	PS029597.D	28 Mar 2025 02:51	AR\AJ	Ok,M
43	PB167233TB	PS029598.D	28 Mar 2025 03:15	AR\AJ	Ok,M
44	I.BLK	PS029599.D	28 Mar 2025 03:39	AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QCBatch ID # PS032825

Review By	Abdul	Review On	3/28/2025 8:31:32 AM
Supervise By	mohammad	Supervise On	4/1/2025 1:29:30 AM
SubDirectory	PS032825	HP Acquire Method	HP Processing Method ps032825 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068		
CCC Internal Standard/PEM	PP24066		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070		

45	HSTDCCC750	PS029600.D	28 Mar 2025 04:27	AR\AJ	Ok,M
46	PB167312BS	PS029601.D	28 Mar 2025 04:51	AR\AJ	Ok,M
47	PB167275TB	PS029602.D	28 Mar 2025 05:15	AR\AJ	Ok,M
48	PB167293BL	PS029603.D	28 Mar 2025 05:39	AR\AJ	Ok,M
49	PB167293BS	PS029604.D	28 Mar 2025 06:03	AR\AJ	Not Ok
50	Q1635-01	PS029605.D	28 Mar 2025 06:27	AR\AJ	Ok,M
51	PB167322BL	PS029606.D	28 Mar 2025 06:51	AR\AJ	Ok,M
52	PB167322BS	PS029607.D	28 Mar 2025 07:15	AR\AJ	Ok,M
53	Q1642-01	PS029608.D	28 Mar 2025 07:39	AR\AJ	Ok,M
54	Q1645-01	PS029609.D	28 Mar 2025 08:03	AR\AJ	Ok,M
55	Q1645-05	PS029610.D	28 Mar 2025 08:27	AR\AJ	Ok,M
56	I.BLK	PS029611.D	28 Mar 2025 08:51	AR\AJ	Ok,M
57	HSTDCCC750	PS029612.D	28 Mar 2025 09:15	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS040225

Review By	Abdul	Review On	4/3/2025 9:25:38 AM
Supervise By	mohammad	Supervise On	4/4/2025 7:53:55 AM
SubDirectory	PS040225	HP Acquire Method	HP Processing Method ps040225 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068		
CCC Internal Standard/PEM	PP24066		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS029655.D	02 Apr 2025 16:20	AR\AJ	Ok
2	I.BLK	PS029656.D	02 Apr 2025 16:44	AR\AJ	Ok
3	HSTDICC200	PS029657.D	02 Apr 2025 17:32	AR\AJ	Ok
4	HSTDICC500	PS029658.D	02 Apr 2025 17:56	AR\AJ	Ok
5	HSTDICC750	PS029659.D	02 Apr 2025 18:44	AR\AJ	Ok,M
6	HSTDICC1000	PS029660.D	02 Apr 2025 19:32	AR\AJ	Ok,M
7	HSTDICC1500	PS029661.D	02 Apr 2025 20:44	AR\AJ	Ok,M
8	HSTDICV750	PS029662.D	02 Apr 2025 21:32	AR\AJ	Ok,M
9	I.BLK	PS029663.D	02 Apr 2025 21:56	AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS040325

Review By	Abdul	Review On	4/3/2025 12:16:42 PM
Supervise By	mohammad	Supervise On	4/5/2025 3:48:57 AM
SubDirectory	PS040325	HP Acquire Method	HP Processing Method ps040225 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068		
CCC Internal Standard/PEM	PP24066		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PS029664.D	03 Apr 2025 08:26	AR\AJ	Ok
2	I.BLK	PS029665.D	03 Apr 2025 08:50	AR\AJ	Ok
3	HSTDCCC750	PS029666.D	03 Apr 2025 09:14	AR\AJ	Ok,M
4	PB167229BS	PS029667.D	03 Apr 2025 09:48	AR\AJ	Not Ok
5	PB167229BSD	PS029668.D	03 Apr 2025 10:12	AR\AJ	Not Ok
6	PB167229BL	PS029669.D	03 Apr 2025 10:36	AR\AJ	Not Ok
7	Q1502-17	PS029670.D	03 Apr 2025 11:06	AR\AJ	Ok,M
8	Q1672-01	PS029671.D	03 Apr 2025 11:30	AR\AJ	Ok,M
9	I.BLK	PS029672.D	03 Apr 2025 12:08	AR\AJ	Ok
10	HSTDCCC750	PS029673.D	03 Apr 2025 12:32	AR\AJ	Ok
11	PB167396BL	PS029674.D	03 Apr 2025 13:18	AR\AJ	Ok,M
12	PB167396BS	PS029675.D	03 Apr 2025 13:42	AR\AJ	Ok,M
13	Q1671-01	PS029676.D	03 Apr 2025 14:06	AR\AJ	Ok,M
14	Q1671-01MS	PS029677.D	03 Apr 2025 14:30	AR\AJ	Ok,M
15	Q1671-01MSD	PS029678.D	03 Apr 2025 14:54	AR\AJ	Ok,M
16	Q1674-01	PS029679.D	03 Apr 2025 15:18	AR\AJ	Ok,M
17	Q1674-03	PS029680.D	03 Apr 2025 15:42	AR\AJ	Ok,M
18	PB167293BS	PS029681.D	03 Apr 2025 16:06	AR\AJ	Ok,M
19	I.BLK	PS029682.D	03 Apr 2025 16:40	AR\AJ	Ok
20	HSTDCCC750	PS029683.D	03 Apr 2025 17:59	AR\AJ	Ok,M
21	PB167433BL	PS029684.D	03 Apr 2025 18:23	AR\AJ	Not Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QCBatch ID # PS040325

Review By	Abdul	Review On	4/3/2025 12:16:42 PM
Supervise By	mohammad	Supervise On	4/5/2025 3:48:57 AM
SubDirectory	PS040325	HP Acquire Method	HP Processing Method ps040225 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068		
CCC Internal Standard/PEM	PP24066		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070		

22	PB167433BS	PS029685.D	03 Apr 2025 18:47	AR\AJ	Not Ok
23	Q1679-01	PS029686.D	03 Apr 2025 19:11	AR\AJ	Not Ok
24	Q1679-05	PS029687.D	03 Apr 2025 19:35	AR\AJ	Not Ok
25	Q1680-01	PS029688.D	03 Apr 2025 19:59	AR\AJ	Not Ok
26	Q1681-01	PS029689.D	03 Apr 2025 20:23	AR\AJ	Not Ok
27	Q1681-01MS	PS029690.D	03 Apr 2025 20:48	AR\AJ	Not Ok
28	Q1681-01MSD	PS029691.D	03 Apr 2025 21:12	AR\AJ	Not Ok
29	I.BLK	PS029692.D	03 Apr 2025 21:36	AR\AJ	Ok
30	HSTDCCC750	PS029693.D	03 Apr 2025 22:00	AR\AJ	Not Ok
31	Q1693-01	PS029694.D	03 Apr 2025 22:48	AR\AJ	Not Ok
32	Q1693-05	PS029695.D	03 Apr 2025 23:12	AR\AJ	Not Ok
33	Q1693-09	PS029696.D	03 Apr 2025 23:36	AR\AJ	Not Ok
34	Q1694-01	PS029697.D	04 Apr 2025 00:00	AR\AJ	Not Ok
35	Q1695-01	PS029698.D	04 Apr 2025 00:24	AR\AJ	Not Ok
36	Q1695-05	PS029699.D	04 Apr 2025 00:48	AR\AJ	Not Ok
37	I.BLK	PS029700.D	04 Apr 2025 01:12	AR\AJ	Ok
38	HSTDCCC750	PS029701.D	04 Apr 2025 01:36	AR\AJ	Not Ok

M : Manual Integration