

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 03/11/2025 03/11/2025

Client Sample No.: CCAL01 **Date Analyzed:** 03/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP070824.D **Time Analyzed:** 09:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.671	5.577	5.777	438.360	500.000	-12.3
Aroclor-1016-2	5.693	5.599	5.799	487.160	500.000	-2.6
Aroclor-1016-3	5.755	5.661	5.861	445.280	500.000	-10.9
Aroclor-1016-4	5.853	5.758	5.958	430.470	500.000	-13.9
Aroclor-1016-5	6.146	6.051	6.251	449.870	500.000	-10.0
Aroclor-1260-1	7.263	7.171	7.371	509.800	500.000	2.0
Aroclor-1260-2	7.517	7.425	7.625	541.690	500.000	8.3
Aroclor-1260-3	7.875	7.783	7.983	537.580	500.000	7.5
Aroclor-1260-4	8.100	8.008	8.208	581.670	500.000	16.3
Aroclor-1260-5	8.421	8.328	8.528	508.500	500.000	1.7
Decachlorobiphenyl	10.239	10.152	10.352	48.460	50.000	-3.1
Tetrachloro-m-xylene	4.519	4.424	4.624	48.130	50.000	-3.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 03/11/2025 03/11/2025

Client Sample No.: CCAL01 **Date Analyzed:** 03/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP070824.D **Time Analyzed:** 09:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.906	4.817	5.017	512.260	500.000	2.5
Aroclor-1016-2	4.926	4.836	5.036	494.660	500.000	-1.1
Aroclor-1016-3	5.103	5.013	5.213	496.150	500.000	-0.8
Aroclor-1016-4	5.144	5.054	5.254	505.180	500.000	1.0
Aroclor-1016-5	5.359	5.270	5.470	508.200	500.000	1.6
Aroclor-1260-1	6.395	6.307	6.507	475.200	500.000	-5.0
Aroclor-1260-2	6.583	6.495	6.695	462.260	500.000	-7.5
Aroclor-1260-3	6.737	6.649	6.849	452.510	500.000	-9.5
Aroclor-1260-4	7.208	7.121	7.321	438.910	500.000	-12.2
Aroclor-1260-5	7.450	7.362	7.562	432.640	500.000	-13.5
Decachlorobiphenyl	8.864	8.781	8.981	40.150	50.000	-19.7
Tetrachloro-m-xylene	3.821	3.729	3.929	49.440	50.000	-1.1