

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 03/11/2025 03/11/2025

Client Sample No.: CCAL02 **Date Analyzed:** 03/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP070835.D **Time Analyzed:** 16:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.674	5.577	5.777	467.990	500.000	-6.4
Aroclor-1016-2	5.695	5.599	5.799	502.150	500.000	0.4
Aroclor-1016-3	5.758	5.661	5.861	470.440	500.000	-5.9
Aroclor-1016-4	5.855	5.758	5.958	460.220	500.000	-8.0
Aroclor-1016-5	6.148	6.051	6.251	479.940	500.000	-4.0
Aroclor-1260-1	7.266	7.171	7.371	557.240	500.000	11.4
Aroclor-1260-2	7.519	7.425	7.625	571.940	500.000	14.4
Aroclor-1260-3	7.877	7.783	7.983	575.750	500.000	15.2
Aroclor-1260-4	8.102	8.008	8.208	529.750	500.000	6.0
Aroclor-1260-5	8.423	8.328	8.528	526.440	500.000	5.3
Decachlorobiphenyl	10.241	10.152	10.352	50.130	50.000	0.3
Tetrachloro-m-xylene	4.521	4.424	4.624	48.950	50.000	-2.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 03/11/2025 03/11/2025

Client Sample No.: CCAL02 **Date Analyzed:** 03/24/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP070835.D **Time Analyzed:** 16:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.909	4.817	5.017	528.990	500.000	5.8
Aroclor-1016-2	4.928	4.836	5.036	511.530	500.000	2.3
Aroclor-1016-3	5.105	5.013	5.213	524.760	500.000	5.0
Aroclor-1016-4	5.146	5.054	5.254	535.160	500.000	7.0
Aroclor-1016-5	5.362	5.270	5.470	521.860	500.000	4.4
Aroclor-1260-1	6.398	6.307	6.507	496.050	500.000	-0.8
Aroclor-1260-2	6.586	6.495	6.695	478.800	500.000	-4.2
Aroclor-1260-3	6.739	6.649	6.849	481.170	500.000	-3.8
Aroclor-1260-4	7.211	7.121	7.321	459.930	500.000	-8.0
Aroclor-1260-5	7.453	7.362	7.562	458.220	500.000	-8.4
Decachlorobiphenyl	8.867	8.781	8.981	42.590	50.000	-14.8
Tetrachloro-m-xylene	3.822	3.729	3.929	51.950	50.000	3.9