

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 03/11/2025 03/11/2025

Client Sample No.: CCAL03 **Date Analyzed:** 03/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP070841.D **Time Analyzed:** 08:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.668	5.577	5.777	467.810	500.000	-6.4
Aroclor-1016-2	5.690	5.599	5.799	526.410	500.000	5.3
Aroclor-1016-3	5.752	5.661	5.861	494.770	500.000	-1.0
Aroclor-1016-4	5.850	5.758	5.958	479.520	500.000	-4.1
Aroclor-1016-5	6.142	6.051	6.251	495.330	500.000	-0.9
Aroclor-1260-1	7.261	7.171	7.371	554.060	500.000	10.8
Aroclor-1260-2	7.513	7.425	7.625	579.910	500.000	16.0
Aroclor-1260-3	7.872	7.783	7.983	561.500	500.000	12.3
Aroclor-1260-4	8.097	8.008	8.208	551.000	500.000	10.2
Aroclor-1260-5	8.418	8.328	8.528	573.180	500.000	14.6
Decachlorobiphenyl	10.237	10.152	10.352	51.300	50.000	2.6
Tetrachloro-m-xylene	4.516	4.424	4.624	50.710	50.000	1.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 03/11/2025 03/11/2025

Client Sample No.: CCAL03 **Date Analyzed:** 03/25/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP070841.D **Time Analyzed:** 08:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.907	4.817	5.017	523.020	500.000	4.6
Aroclor-1016-2	4.926	4.836	5.036	532.400	500.000	6.5
Aroclor-1016-3	5.103	5.013	5.213	541.040	500.000	8.2
Aroclor-1016-4	5.146	5.054	5.254	476.330	500.000	-4.7
Aroclor-1016-5	5.359	5.270	5.470	562.570	500.000	12.5
Aroclor-1260-1	6.395	6.307	6.507	522.520	500.000	4.5
Aroclor-1260-2	6.583	6.495	6.695	506.540	500.000	1.3
Aroclor-1260-3	6.737	6.649	6.849	492.950	500.000	-1.4
Aroclor-1260-4	7.209	7.121	7.321	466.890	500.000	-6.6
Aroclor-1260-5	7.450	7.362	7.562	457.580	500.000	-8.5
Decachlorobiphenyl	8.865	8.781	8.981	42.300	50.000	-15.4
Tetrachloro-m-xylene	3.821	3.729	3.929	54.550	50.000	9.1