

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>WALS01</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q1626</u>
Instrument ID:	<u>ECD_L</u>	Calibration Date(s):	<u>03/11/2025</u> <u>03/11/2025</u>
		Calibration Times:	<u>11:43</u> <u>12:37</u>

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PL094574.D</u>	RT 750 = <u>PL094575.D</u>
RT 500 = <u>PL094576.D</u>	RT 250 = <u>PL094577.D</u>	RT 050 = <u>PL094578.D</u>

COMPOUND	RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
							FROM	TO
Decachlorobiphenyl	9.06	9.06	9.06	9.06	9.06	9.06	8.96	9.16
Tetrachloro-m-xylene	3.54	3.54	3.54	3.54	3.54	3.54	3.44	3.64
Decachlorobiphenyl	9.06	9.06	9.05	9.06	9.06	9.06	8.96	9.16
Tetrachloro-m-xylene	3.54	3.54	3.54	3.54	3.54	3.54	3.44	3.64
Toxaphene-1 (1)	6.24	6.24	6.24	6.24	6.23	6.24	6.14	6.34
Toxaphene-2 (2)	6.44	6.44	6.44	6.44	6.44	6.44	6.34	6.54
Toxaphene-3 (3)	7.06	7.06	7.06	7.06	7.06	7.06	6.96	7.16
Toxaphene-4 (4)	7.15	7.15	7.15	7.15	7.15	7.15	7.05	7.25
Toxaphene-5 (5)	7.93	7.93	7.93	7.93	7.94	7.93	7.83	8.03

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name: Alliance

Contract: **WALS01**

Lab Code: ACE

SDG NO.: Q1626

Instrument ID: ECD_L

Calibration Date(s): **03/11/2025** **03/11/2025**

Calibration Times: 11:43 12:37

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 =	<u>PL094574.D</u>	RT 750 =	<u>PL094575.D</u>
	RT 500 =	PL094576.D	RT 250 =	PL094577.D
			RT 050 =	PL094578.D

[illegible]