

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 03/11/2025 03/11/2025

Client Sample No.: CCAL01 **Date Analyzed:** 03/25/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL094832.D **Time Analyzed:** 11:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.708	6.610	6.810	53.660	50.000	7.3
4,4'-DDE	6.191	6.093	6.293	52.030	50.000	4.1
4,4'-DDT	7.023	6.924	7.124	52.040	50.000	4.1
Aldrin	5.255	5.156	5.356	45.500	50.000	-9.0
alpha-BHC	3.993	3.894	4.094	47.540	50.000	-4.9
alpha-Chlordane	6.017	5.918	6.118	47.280	50.000	-5.4
beta-BHC	4.525	4.425	4.625	46.330	50.000	-7.3
Decachlorobiphenyl	9.054	8.956	9.156	50.670	50.000	1.3
delta-BHC	4.771	4.673	4.873	47.760	50.000	-4.5
Dieldrin	6.342	6.244	6.444	47.880	50.000	-4.2
Endosulfan I	6.068	5.969	6.169	47.310	50.000	-5.4
Endosulfan II	6.793	6.694	6.894	48.260	50.000	-3.5
Endosulfan sulfate	7.158	7.059	7.259	48.260	50.000	-3.5
Endrin	6.571	6.474	6.674	46.840	50.000	-6.3
Endrin aldehyde	6.923	6.824	7.024	48.520	50.000	-3.0
Endrin ketone	7.643	7.544	7.744	49.010	50.000	-2.0
gamma-BHC (Lindane)	4.326	4.227	4.427	47.530	50.000	-4.9
gamma-Chlordane	5.938	5.840	6.040	46.950	50.000	-6.1
Heptachlor	4.913	4.815	5.015	45.980	50.000	-8.0
Heptachlor epoxide	5.681	5.583	5.783	46.860	50.000	-6.3
Methoxychlor	7.500	7.400	7.600	56.460	50.000	12.9
Tetrachloro-m-xylene	3.537	3.438	3.638	47.620	50.000	-4.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 03/11/2025 03/11/2025

Client Sample No.: CCAL01 **Date Analyzed:** 03/25/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL094832.D **Time Analyzed:** 11:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.778	5.682	5.882	54.800	50.000	9.6
4,4'-DDE	5.222	5.127	5.327	52.740	50.000	5.5
4,4'-DDT	6.028	5.932	6.132	54.820	50.000	9.6
Aldrin	4.218	4.122	4.322	50.820	50.000	1.6
alpha-BHC	3.271	3.174	3.374	52.360	50.000	4.7
alpha-Chlordane	5.034	4.938	5.138	50.280	50.000	0.6
beta-BHC	3.901	3.805	4.005	51.660	50.000	3.3
Decachlorobiphenyl	7.904	7.807	8.007	50.180	50.000	0.4
delta-BHC	4.130	4.033	4.233	52.390	50.000	4.8
Dieldrin	5.353	5.258	5.458	50.550	50.000	1.1
Endosulfan I	5.090	4.994	5.194	42.410	50.000	-15.2
Endosulfan II	5.925	5.829	6.029	50.470	50.000	0.9
Endosulfan sulfate	6.328	6.231	6.431	51.470	50.000	2.9
Endrin	5.630	5.534	5.734	50.820	50.000	1.6
Endrin aldehyde	6.105	6.008	6.208	50.100	50.000	0.2
Endrin ketone	6.832	6.736	6.936	53.330	50.000	6.7
gamma-BHC (Lindane)	3.601	3.504	3.704	51.500	50.000	3.0
gamma-Chlordane	4.971	4.874	5.074	50.610	50.000	1.2
Heptachlor	3.939	3.842	4.042	50.290	50.000	0.6
Heptachlor epoxide	4.721	4.625	4.825	50.460	50.000	0.9
Methoxychlor	6.604	6.507	6.707	56.290	50.000	12.6
Tetrachloro-m-xylene	2.769	2.672	2.872	51.730	50.000	3.5