

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q1626</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Walsh Construction Company II, LLC</u>	DataFile :	<u>PL094838.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec Qual	RPD		Low	Limits	
			Result					Qual	RPD		High	RPD
Lab Sample ID:	Q1626-01MSD	Client Sample ID:		CO-32-1MSD								
	(Column 1)											
	alpha-BHC	17.57	0	15.1	ug/kg	90		1		60	144	20
	beta-BHC	17.57	0	14.5	ug/kg	107		2		54	143	20
	delta-BHC	17.57	0	25.8	ug/kg	147		4		29	151	20
	gamma-BHC (Lindane)	17.57	0	17.7	ug/kg	113		5		61	140	20
	Heptachlor	17.57	0	15.8	ug/kg	90		4		63	135	20
	Aldrin	17.57	0	15.4	ug/kg	105		4		49	139	20
	Heptachlor epoxide	17.57	0	28.3	ug/kg	161	*	5		41	156	20
	Endosulfan I	17.57	0	15.4	ug/kg	88		2		56	142	20
	Dieldrin	17.57	0	20.6	ug/kg	233	*	3		47	161	20
	4,4'-DDE	17.57	0	23.8	ug/kg	135		3		55	136	20
	Endrin	17.57	0	29.2	ug/kg	166	*	3		57	139	20
	Endosulfan II	17.57	0	16.8	ug/kg	131		0		40	163	20
	4,4'-DDD	17.57	0	50.5	ug/kg	287	*	4		47	163	20
	Endosulfan sulfate	17.57	0	23.8	ug/kg	135		4		62	139	20
	4,4'-DDT	17.57	0	21.6	ug/kg	302	*	2		51	146	20
	Methoxychlor	17.57	0	15.0	ug/kg	85		19		54	136	20
	Endrin ketone	17.57	0	16.7	ug/kg	95		2		60	129	20
	Endrin aldehyde	17.57	0	11.8	ug/kg	81		9		59	132	20
	alpha-Chlordane	17.57	11.1	25.6	ug/kg	93		1		39	166	20
	gamma-Chlordane	17.57	11.1	32.0	ug/kg	119		5		44	175	20
Lab Sample ID:	Q1626-01MSD	Client Sample ID:		CO-32-1MSD								
	(Column 2)											
	alpha-BHC	17.57	0	15.8	ug/kg	0		2		60	144	20
	beta-BHC	17.57	0	18.8	ug/kg	0		2		54	143	20
	delta-BHC	17.57	0	13.8	ug/kg	0		3		29	151	20
	gamma-BHC (Lindane)	17.57	0	19.8	ug/kg	0		5		61	140	20
	Heptachlor	17.57	0	15.8	ug/kg	0		2		63	135	20
	Aldrin	17.57	0	18.5	ug/kg	0		4		49	139	20
	Heptachlor epoxide	17.57	0	16.5	ug/kg	0		1		41	156	20
	Endosulfan I	17.57	0	10.0	ug/kg	0		2		56	142	20
	Dieldrin	17.57	0	40.9	ug/kg	0		3		47	161	20
	4,4'-DDE	17.57	0	18.5	ug/kg	0		4		55	136	20
	Endrin	17.57	0	20.1	ug/kg	0		3		57	139	20
	Endosulfan II	17.57	0	23.1	ug/kg	0		0		40	163	20
	4,4'-DDD	17.57	0	22.1	ug/kg	0		1		47	163	20
	Endosulfan sulfate	17.57	0	21.9	ug/kg	0		9		62	139	20

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q1626</u>	Analytical Method:	<u>8081B</u>
Client:	<u>Walsh Construction Company II, LLC</u>	DataFile :	<u>PL094838.D</u>

Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
		Result	Result			Qual	RPD	Qual	Low	High	RPD
4,4'-DDT	17.57	0	53.0	ug/kg	0		2		51	146	20
Methoxychlor	17.57	0	8.00	ug/kg	0		23		54	136	20
Endrin ketone	17.57	0	11.7	ug/kg	0		13		60	129	20
Endrin aldehyde	17.57	0	14.3	ug/kg	0		9		59	132	20
alpha-Chlordane	17.57	9.6	27.4	ug/kg	0		1		39	166	20
gamma-Chlordane	17.57	8.2	27.3	ug/kg	0		1		44	175	20