

Continuing Calibration Report for SequenceID : FE032425AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE052939.D

Aliphatic C9-C12	9562712.000	60.000	3.147	6.791	159378.533	155948.436	-2.200
Aliphatic C12-C16	6277985.000	40.000	6.792	10.244	156949.625	152098.489	-3.189
Aliphatic C16-C21	8955302.000	60.000	10.245	13.623	149255.033	146548.085	-1.847
Aliphatic C21-C28	11091377.000	80.000	13.624	17.297	138642.213	142451.851	2.674
Aliphatic C28-C40	14513964.000	120.000	17.298	22.225	120949.700	127589.108	5.204
Aliphatic EPH	50401340.000	360.000	3.147	22.225	140003.722	141501.588	1.059