

Continuing Calibration Report for SequenceID : FE032425AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FE052950.D

Aliphatic C9-C12	9782151.000	60.000	3.147	6.791	163035.850	155948.436	-4.545
Aliphatic C12-C16	6351173.000	40.000	6.792	10.244	158779.325	152098.489	-4.392
Aliphatic C16-C21	8991010.000	60.000	10.245	13.623	149850.167	146548.085	-2.253
Aliphatic C21-C28	11184998.000	80.000	13.624	17.297	139812.475	142451.851	1.853
Aliphatic C28-C40	15005050.000	120.000	17.298	22.225	125042.083	127589.108	1.996
Aliphatic EPH	51314382.000	360.000	3.147	22.225	142539.950	141501.588	-0.734