

SOLID EPH_NF LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Walsh Construction Company II, LLC
Lab Code: CHEM **Cas No:** Q1626 **SAS No :** Q1626 **SDG No:** Q1626
Sample No : PB167270BS **Datafile:** FE052941.D
Client ID : PB167270BS

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aliphatic C28-C40	30.0	38.4	128		(40-140)
Aliphatic C9-C28	99.9	105	106		(40-140)