

SOLID EPH_NF LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** Walsh Construction Company II, LLC
Lab Code: CHEM **Cas No:** Q1626 **SAS No :** Q1626 **SDG No:** Q1626
Sample No : PB167270BSD **Datafile:** FE052942.D
Client ID : PB167270BSD

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aliphatic C28-C40	30.0	41.0	136		6.6 (40-140)	50
Aliphatic C9-C28	100.1	111	112		6 (40-140)	50