

SOIL GASOLINE RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATION

Lab Name: Chemtech **Client:** Walsh Construction Company II, LLC
Lab Code: CHEM **Cas No:** Q1626 **SAS No :** Q1626 **SDG No:** Q1626
Matrix Spike - EPA Sample No : BSF0324S2 **Datafile:** FB031613.D

COMPOUND	SPIKE ADDED ug/kg	CONCENTRATION ug/kg	LCS/LCSD CONCENTRATION ug/kg	% REC	QC LIMITS
GRO	180	0	203	113	50-150

LCS/LCSD % Recovery RPD : 1.0