

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q1626

Continuing Calib Date: 03/26/2025

Initial Calibration Date(s): 03/26/2025

03/26/2025

Continuing Calib Time: 16:54

Initial Calibration Time(s): 12:36

15:28

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.30	7.30	7.20	7.40	0.00
2,4-DCAA	7.12	7.12	7.02	7.22	0.00
DICHLORPROP	8.00	8.00	7.90	8.10	0.00
2,4-D	8.23	8.23	8.13	8.33	0.00
2,4,5-TP(Silvex)	9.09	9.09	8.99	9.19	0.00
2,4,5-T	9.38	9.38	9.28	9.48	0.00
2,4-DB	9.95	9.95	9.85	10.05	0.00
Dinoseb	11.13	11.13	11.03	11.23	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: WALS01

Lab Code: ACE

SDG NO.: Q1626

Continuing Calib Date: 03/26/2025

Initial Calibration Date(s): 03/26/2025

03/26/2025

Continuing Calib Time: 16:54

Initial Calibration Time(s): 12:36

15:28

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.73	7.73	7.63	7.83	0.00
2,4-DCAA	7.54	7.54	7.44	7.64	0.00
DICHLORPROP	8.43	8.43	8.33	8.53	0.00
2,4-D	8.76	8.75	8.65	8.85	-0.01
2,4,5-TP(Silvex)	9.64	9.64	9.54	9.74	0.00
2,4,5-T	10.05	10.05	9.95	10.15	0.00
2,4-DB	10.61	10.61	10.51	10.71	0.00
Dinoseb	10.98	10.98	10.88	11.08	0.00