

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1626 SAS No.: Q1626 SDG NO.: Q1626

Continuing Calib Date: 03/27/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 13:13 Initial Calibration Time(s): 10:13 11:49

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.29	7.29	7.19	7.39	0.00
2,4-DCAA	7.12	7.12	7.02	7.22	0.01
DICHLORPROP	7.99	7.99	7.89	8.09	0.00
2,4-D	8.22	8.22	8.12	8.32	0.00
2,4,5-TP(Silvex)	9.08	9.08	8.98	9.18	0.00
2,4,5-T	9.37	9.37	9.27	9.47	0.00
2,4-DB	9.94	9.94	9.84	10.04	0.00
Dinoseb	11.13	11.13	11.03	11.23	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1626 SAS No.: Q1626 SDG NO.: Q1626

Continuing Calib Date: 03/27/2025 Initial Calibration Date(s): 03/27/2025 03/27/2025

Continuing Calib Time: 13:13 Initial Calibration Time(s): 10:13 11:49

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.73	7.73	7.63	7.83	0.01
2,4-DCAA	7.54	7.54	7.44	7.64	0.01
DICHLORPROP	8.43	8.43	8.33	8.53	0.00
2,4-D	8.75	8.75	8.65	8.85	0.00
2,4,5-TP(Silvex)	9.63	9.63	9.53	9.73	0.00
2,4,5-T	10.04	10.04	9.94	10.14	0.00
2,4-DB	10.60	10.60	10.50	10.70	0.00
Dinoseb	10.97	10.97	10.87	11.07	0.00