

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1626 SAS No.: Q1626 SDG NO.: Q1626

Continuing Calib Date: 04/03/2025 Initial Calibration Date(s): 04/02/2025 04/02/2025

Continuing Calib Time: 09:14 Initial Calibration Time(s): 17:32 20:44

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.14	7.14	7.04	7.24	0.00
2,4-DCAA	6.96	6.96	6.86	7.06	0.00
DICHLORPROP	7.82	7.82	7.72	7.92	0.00
2,4-D	8.04	8.04	7.94	8.14	0.00
2,4,5-TP(Silvex)	8.89	8.89	8.79	8.99	0.00
2,4,5-T	9.17	9.17	9.07	9.27	0.00
2,4-DB	9.73	9.73	9.63	9.83	0.00
Dinoseb	10.89	10.90	10.80	11.00	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1626 SAS No.: Q1626 SDG NO.: Q1626

Continuing Calib Date: 04/03/2025 Initial Calibration Date(s): 04/02/2025 04/02/2025

Continuing Calib Time: 09:14 Initial Calibration Time(s): 17:32 20:44

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
DICAMBA	7.66	7.66	7.56	7.76	0.00
2,4-DCAA	7.48	7.48	7.38	7.58	0.01
DICHLORPROP	8.36	8.36	8.26	8.46	0.00
2,4-D	8.67	8.67	8.57	8.77	0.00
2,4,5-TP(Silvex)	9.55	9.55	9.45	9.65	0.00
2,4,5-T	9.96	9.96	9.86	10.06	0.00
2,4-DB	10.51	10.51	10.41	10.61	0.00
Dinoseb	10.89	10.89	10.79	10.99	0.00