

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1626 SAS No.: Q1626 SDG NO.: Q1626

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL02 Date Analyzed: 03/28/2025

Lab Sample No.: HSTDCCC750 Data File : PS029600.D Time Analyzed: 04:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.373	9.273	9.473	751.370	712.500	5.5
2,4,5-TP(Silvex)	9.082	8.982	9.182	707.180	712.500	-0.7
2,4-D	8.221	8.122	8.322	728.240	705.000	3.3
2,4-DB	9.942	9.843	10.043	818.970	712.500	14.9
2,4-DCAA	7.116	7.015	7.215	733.820	750.000	-2.2
DICAMBA	7.295	7.194	7.394	683.590	705.000	-3.0
DICHLORPROP	7.992	7.891	8.091	687.210	705.000	-2.5
Dinoseb	11.127	11.026	11.226	675.180	705.000	-4.2

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1626 SAS No.: Q1626 SDG NO.: Q1626

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL02 Date Analyzed: 03/28/2025

Lab Sample No.: HSTDCCC750 Data File : PS029600.D Time Analyzed: 04:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.040	9.943	10.143	707.300	712.500	-0.7
2,4,5-TP(Silvex)	9.630	9.534	9.734	710.210	712.500	-0.3
2,4-D	8.745	8.648	8.848	692.900	705.000	-1.7
2,4-DB	10.598	10.502	10.702	637.690	712.500	-10.5
2,4-DCAA	7.535	7.437	7.637	732.700	750.000	-2.3
DICAMBA	7.724	7.626	7.826	703.150	705.000	-0.3
DICHLORPROP	8.426	8.328	8.528	672.740	705.000	-4.6
Dinoseb	10.971	10.874	11.074	674.270	705.000	-4.4