

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032825\
 Data File : PS029588.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Mar 2025 22:51
 Operator : AR\AJ
 Sample : HSTDCCC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDCCC750

Manual Integrations
 APPROVED

Reviewed By : Abdul Mirza 03/28/2025
 Supervised By : mohammad ahmed 04/01/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 28 03:24:08 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032825.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 27 12:24:58 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.114	7.534	2855.1E6	650.1E6	743.506	753.071
Target Compounds							
1) T	Dalapon	2.593	2.605	4052.5E6	1379.7E6	683.564	665.121
2) T	3,5-DICHL...	6.304	6.523	4014.6E6	868.4E6	702.275	704.799
3) T	4-Nitroph...	6.919	7.070	937.6E6	561.4E6	881.247	694.985
5) T	DICAMBA	7.293	7.724	10969.2E6	3458.0E6	691.938	723.307
6) T	MCP	7.474	7.831	822.9E6	150.4E6	72.294	72.659
7) T	MCPA	7.620	8.066	1038.4E6	202.6E6	72.201	70.495
8) T	DICHLORPROP	7.990	8.425	3043.9E6	860.5E6	693.066	690.808
9) T	2,4-D	8.220	8.745	2992.8E6	976.1E6	741.929	714.596
10) T	Pentachlo...	8.510	9.254	44453.7E6	16328.1E6	748.127	735.674
11) T	2,4,5-TP ...	9.080	9.630	16735.2E6	6431.8E6	718.403	728.837
12) T	2,4,5-T	9.371	10.040	15680.7E6	5827.1E6	761.262	731.071
13) T	2,4-DB	9.939	10.598	1543.1E6	604.2E6	808.180m	657.449
14) T	DINOSEB	11.124	10.971	12519.5E6	4192.9E6	706.115	707.317
15) T	Picloram	10.947	12.032	12846.5E6	9254.2E6	892.519	741.752
16) T	DCPA	11.422	11.997	20726.8E6	8317.5E6	718.910	741.856

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032825\
Data File : PS029588.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 22:51
Operator : AR\AJ
Sample : HSTDCCC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDCCC750

Manual Integrations
APPROVED

Reviewed By : Abdul Mirza 03/28/2025
Supervised By : mohammad ahmed 04/01/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 28 03:24:08 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032825.M
Quant Title : 8080.M
QLast Update : Thu Mar 27 12:24:58 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

