

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS032825

Review By	Abdul	Review On	3/28/2025 8:31:32 AM
Supervise By	mohammad	Supervise On	4/1/2025 1:29:30 AM
SubDirectory	PS032825	HP Acquire Method	HP Processing Method ps032825 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM	PP24066
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS029556.D	27 Mar 2025 09:25		AR\AJ	Ok
2	I.BLK	I.BLK	PS029557.D	27 Mar 2025 09:49		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS029558.D	27 Mar 2025 10:13		AR\AJ	Ok
4	HSTDICC500	HSTDICC500	PS029559.D	27 Mar 2025 10:37		AR\AJ	Ok
5	HSTDICC750	HSTDICC750	PS029560.D	27 Mar 2025 11:01	Method Fail for comp#10 and 15	AR\AJ	Ok
6	HSTDICC1000	HSTDICC1000	PS029561.D	27 Mar 2025 11:25		AR\AJ	Ok
7	HSTDICC1500	HSTDICC1500	PS029562.D	27 Mar 2025 11:49		AR\AJ	Ok
8	HSTDICV750	ICVPS032825	PS029563.D	27 Mar 2025 12:25	Method Fail for comp#10 and 15	AR\AJ	Ok
9	I.BLK	I.BLK	PS029564.D	27 Mar 2025 12:49		AR\AJ	Ok
10	HSTDCCC750	HSTDCCC750	PS029565.D	27 Mar 2025 13:13		AR\AJ	Ok
11	Q1609-03	WC-SCRN-01-C	PS029566.D	27 Mar 2025 13:37		AR\AJ	Ok,M
12	Q1625-01	FRAC-TANK-A1291	PS029567.D	27 Mar 2025 14:01		AR\AJ	Ok
13	Q1632-02	PIER-1-2	PS029568.D	27 Mar 2025 14:25		AR\AJ	Ok
14	Q1630-01	VNJ-206	PS029569.D	27 Mar 2025 14:49		AR\AJ	Ok,M
15	Q1636-01	WC-1	PS029570.D	27 Mar 2025 15:13		AR\AJ	Ok,M
16	Q1636-05	WC-2	PS029571.D	27 Mar 2025 15:38		AR\AJ	Ok,M
17	Q1636-09	WC-3	PS029572.D	27 Mar 2025 16:02		AR\AJ	Ok
18	Q1636-13	WC-4	PS029573.D	27 Mar 2025 16:26		AR\AJ	Ok

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS032825

Review By	Abdul	Review On	3/28/2025 8:31:32 AM
Supervise By	mohammad	Supervise On	4/1/2025 1:29:30 AM
SubDirectory	PS032825	HP Acquire Method	HP Processing Method ps032825 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068		
CCC Internal Standard/PEM	PP24066		
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070		

19	Q1636-17	WC-5	PS029574.D	27 Mar 2025 16:50		AR\AJ	Ok,M
20	I.BLK	I.BLK	PS029575.D	27 Mar 2025 17:14		AR\AJ	Ok,M
21	HSTDCCC750	HSTDCCC750	PS029576.D	27 Mar 2025 18:02		AR\AJ	Ok,M
22	Q1609-03MS	WC-SCRN-01-CMS	PS029577.D	27 Mar 2025 18:26	Comp#7 recovery fail	AR\AJ	Ok,M
23	Q1609-03MSD	WC-SCRN-01-CMSD	PS029578.D	27 Mar 2025 18:50		AR\AJ	Ok,M
24	Q1636-05MS	WC-2MS	PS029579.D	27 Mar 2025 19:14	Comp#6,10 recovery fail	AR\AJ	Ok,M
25	Q1636-05MSD	WC-2MSD	PS029580.D	27 Mar 2025 19:38	Comp#3,6,10 recovery fail	AR\AJ	Ok,M
26	Q1648-01	TP-4	PS029581.D	27 Mar 2025 20:02		AR\AJ	Ok
27	Q1648-03	TP-5	PS029582.D	27 Mar 2025 20:26		AR\AJ	Ok,M
28	Q1648-05	TP-3	PS029583.D	27 Mar 2025 20:50		AR\AJ	Ok,M
29	Q1648-07	TP-6	PS029584.D	27 Mar 2025 21:14		AR\AJ	Ok,M
30	Q1648-09	TP-7	PS029585.D	27 Mar 2025 21:38		AR\AJ	Ok,M
31	Q1648-11	TP-8	PS029586.D	27 Mar 2025 22:02		AR\AJ	Ok,M
32	I.BLK	I.BLK	PS029587.D	27 Mar 2025 22:26		AR\AJ	Ok,M
33	HSTDCCC750	HSTDCCC750	PS029588.D	27 Mar 2025 22:51	Comp#3,15 high in 1st column	AR\AJ	Ok,M
34	Q1648-13	TP-2	PS029589.D	27 Mar 2025 23:39		AR\AJ	Ok,M
35	Q1648-13MS	TP-2MS	PS029590.D	28 Mar 2025 00:03	some compound recovery fail	AR\AJ	Ok,M
36	Q1648-13MSD	TP-2MSD	PS029591.D	28 Mar 2025 00:27	some compound recovery fail	AR\AJ	Ok,M
37	Q1627-01	GRID-LINE-1.2-NORTH	PS029592.D	28 Mar 2025 00:51		AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS032825

Review By	Abdul	Review On	3/28/2025 8:31:32 AM
Supervise By	mohammad	Supervise On	4/1/2025 1:29:30 AM
SubDirectory	PS032825	HP Acquire Method	HP Processing Method ps032825 8151
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068		
CCC	PP24066		
Internal Standard/PEM	PP24069,PP24070		
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	Q1626-01	CO-32-1	PS029593.D	28 Mar 2025 01:15		AR\AJ	Ok,M
39	Q1626-01MS	CO-32-1MS	PS029594.D	28 Mar 2025 01:39	some compound recovery fail	AR\AJ	Ok,M
40	Q1626-01MSD	CO-32-1MSD	PS029595.D	28 Mar 2025 02:03	some compound recovery fail	AR\AJ	Ok,M
41	Q1626-03	CO-32-1	PS029596.D	28 Mar 2025 02:27		AR\AJ	Ok,M
42	PB167312BL	PB167312BL	PS029597.D	28 Mar 2025 02:51		AR\AJ	Ok,M
43	PB167233TB	PB167233TB	PS029598.D	28 Mar 2025 03:15		AR\AJ	Ok,M
44	I.BLK	I.BLK	PS029599.D	28 Mar 2025 03:39		AR\AJ	Ok,M
45	HSTDCCC750	HSTDCCC750	PS029600.D	28 Mar 2025 04:27	Comp#3,15 high in 1st column	AR\AJ	Ok,M
46	PB167312BS	PB167312BS	PS029601.D	28 Mar 2025 04:51		AR\AJ	Ok,M
47	PB167275TB	PB167275TB	PS029602.D	28 Mar 2025 05:15		AR\AJ	Ok,M
48	PB167293BL	PB167293BL	PS029603.D	28 Mar 2025 05:39		AR\AJ	Ok,M
49	PB167293BS	PB167293BS	PS029604.D	28 Mar 2025 06:03	2,4-DCAA high in both column , Almost all compound recovery fail	AR\AJ	Not Ok
50	Q1635-01	TP-2	PS029605.D	28 Mar 2025 06:27		AR\AJ	Ok,M
51	PB167322BL	PB167322BL	PS029606.D	28 Mar 2025 06:51		AR\AJ	Ok,M
52	PB167322BS	PB167322BS	PS029607.D	28 Mar 2025 07:15		AR\AJ	Ok,M
53	Q1642-01	PAT-1-032525	PS029608.D	28 Mar 2025 07:39		AR\AJ	Ok,M
54	Q1645-01	TP-4	PS029609.D	28 Mar 2025 08:03		AR\AJ	Ok,M
55	Q1645-05	TP-1	PS029610.D	28 Mar 2025 08:27		AR\AJ	Ok,M

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS032825

Review By	Abdul	Review On	3/28/2025 8:31:32 AM
Supervise By	mohammad	Supervise On	4/1/2025 1:29:30 AM
SubDirectory	PS032825	HP Acquire Method	HP Processing Method ps032825 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM	PP24066
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070

56	I.BLK	I.BLK	PS029611.D	28 Mar 2025 08:51		AR\AJ	Ok,M
57	HSTDCCC750	HSTDCCC750	PS029612.D	28 Mar 2025 09:15	Comp#3,13,15 high in 1st column	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS040225

Review By	Abdul	Review On	4/3/2025 9:25:38 AM
Supervise By		Supervise On	
SubDirectory	PS040225	HP Acquire Method	HP Processing Method ps040225 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM	PP24066
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS029655.D	02 Apr 2025 16:20		AR\AJ	Ok
2	I.BLK	I.BLK	PS029656.D	02 Apr 2025 16:44		AR\AJ	Ok
3	HSTDICC200	HSTDICC200	PS029657.D	02 Apr 2025 17:32		AR\AJ	Ok
4	HSTDICC500	HSTDICC500	PS029658.D	02 Apr 2025 17:56		AR\AJ	Ok
5	HSTDICC750	HSTDICC750	PS029659.D	02 Apr 2025 18:44		AR\AJ	Ok,NS
6	HSTDICC1000	HSTDICC1000	PS029660.D	02 Apr 2025 19:32		AR\AJ	Ok,NS
7	HSTDICC1500	HSTDICC1500	PS029661.D	02 Apr 2025 20:44		AR\AJ	Ok,NS
8	HSTDICV750	ICVPS040225	PS029662.D	02 Apr 2025 21:32		AR\AJ	Ok,NS
9	I.BLK	I.BLK	PS029663.D	02 Apr 2025 21:56		AR\AJ	Ok

M : Manual Integration

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS040325

Review By	Abdul	Review On	4/3/2025 12:16:42 PM
Supervise By		Supervise On	
SubDirectory	PS040325	HP Acquire Method	HP Processing Method ps040225 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM	PP24066
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PS029664.D	03 Apr 2025 08:26		AR\AJ	Ok
2	I.BLK	I.BLK	PS029665.D	03 Apr 2025 08:50		AR\AJ	Ok
3	HSTDCCC750	HSTDCCC750	PS029666.D	03 Apr 2025 09:14		AR\AJ	Ok,NS
4	PB167229BS	PB167229BS	PS029667.D	03 Apr 2025 09:48		AR\AJ	Ok,NS
5	PB167229BSD	PB167229BSD	PS029668.D	03 Apr 2025 10:12		AR\AJ	Ok,NS
6	PB167229BL	PB167229BL	PS029669.D	03 Apr 2025 10:36		AR\AJ	Ok
7	Q1502-17	PT-HERB-WP	PS029670.D	03 Apr 2025 11:06		AR\AJ	Ok,NS
8	Q1672-01	TP-8	PS029671.D	03 Apr 2025 11:30		AR\AJ	Ok,NS
9	I.BLK	I.BLK	PS029672.D	03 Apr 2025 12:08		AR\AJ	Ok
10	HSTDCCC750	HSTDCCC750	PS029673.D	03 Apr 2025 12:32		AR\AJ	Ok
11	PB167396BL	PB167396BL	PS029674.D	03 Apr 2025 13:18		AR\AJ	Ok,NR
12	PB167396BS	PB167396BS	PS029675.D	03 Apr 2025 13:42		AR\AJ	Ok,NR
13	Q1671-01	WC-1	PS029676.D	03 Apr 2025 14:06		AR\AJ	Ok,NR
14	Q1671-01MS	WC-1MS	PS029677.D	03 Apr 2025 14:30	some compound recovery fail	AR\AJ	Ok,NR
15	Q1671-01MSD	WC-1MSD	PS029678.D	03 Apr 2025 14:54	some compound recovery fail	AR\AJ	Ok,NR
16	Q1674-01	RT5358	PS029679.D	03 Apr 2025 15:18		AR\AJ	Ok,NR
17	Q1674-03	72-11991	PS029680.D	03 Apr 2025 15:42		AR\AJ	Ok,NR
18	PB167293BS	PB167293BS	PS029681.D	03 Apr 2025 16:06		AR\AJ	Ok,NR

Instrument ID: ECD_S

Daily Analysis Runlog For Sequence/QC Batch ID # PS040325

Review By	Abdul	Review On	4/3/2025 12:16:42 PM
Supervise By		Supervise On	
SubDirectory	PS040325	HP Acquire Method	HP Processing Method ps040225 8151

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24064,PP24065,PP24066,PP24067,PP24068
CCC Internal Standard/PEM	PP24066
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24069,PP24070

19	I.BLK	I.BLK	PS029682.D	03 Apr 2025 16:40		AR\AJ	Ok
20	HSTDCCC750	HSTDCCC750	PS029683.D	03 Apr 2025 17:59		AR\AJ	Ok,NR

M : Manual Integration