



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** Q1626

**Client:** Walsh Construction Company II, LLC

**Analytical Method:** 8151A **Datafile :** PS029681.D

| Lab Sample ID | Parameter        | Spike | Result | Units | Rec | RPD | Qual | RPD  |     | Limits |     |
|---------------|------------------|-------|--------|-------|-----|-----|------|------|-----|--------|-----|
|               |                  |       |        |       |     |     |      | Qual | Low | High   | RPD |
| PB167293BS    | DICAMBA          | 166.6 | 187    | ug/Kg | 112 |     |      |      | 72  | 129    |     |
|               | DICHLORPROP      | 166.6 | 184    | ug/Kg | 110 |     |      |      | 77  | 135    |     |
|               | 2,4-D            | 166.6 | 186    | ug/Kg | 112 |     |      |      | 65  | 144    |     |
|               | 2,4,5-TP(Silvex) | 166.6 | 192    | ug/Kg | 115 |     |      |      | 74  | 146    |     |
|               | 2,4,5-T          | 166.6 | 194    | ug/Kg | 116 |     |      |      | 77  | 134    |     |
|               | 2,4-DB           | 166.6 | 188    | ug/Kg | 113 |     |      |      | 72  | 122    |     |
|               | Dinoseb          | 166.6 | 188    | ug/Kg | 113 |     |      |      | 74  | 132    |     |