

### CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1626 SAS No.: Q1626 SDG NO.: Q1626

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/24/2025

Lab Sample No.: AR1660CCC500 Data File : PP070824.D Time Analyzed: 09:06

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|--------|-----------|--------|--------------------|-------------------|-------|
|                      |        | FROM      | TO     |                    |                   |       |
| Aroclor-1016-1       | 5.671  | 5.577     | 5.777  | 438.360            | 500.000           | -12.3 |
| Aroclor-1016-2       | 5.693  | 5.599     | 5.799  | 487.160            | 500.000           | -2.6  |
| Aroclor-1016-3       | 5.755  | 5.661     | 5.861  | 445.280            | 500.000           | -10.9 |
| Aroclor-1016-4       | 5.853  | 5.758     | 5.958  | 430.470            | 500.000           | -13.9 |
| Aroclor-1016-5       | 6.146  | 6.051     | 6.251  | 449.870            | 500.000           | -10.0 |
| Aroclor-1260-1       | 7.263  | 7.171     | 7.371  | 509.800            | 500.000           | 2.0   |
| Aroclor-1260-2       | 7.517  | 7.425     | 7.625  | 541.690            | 500.000           | 8.3   |
| Aroclor-1260-3       | 7.875  | 7.783     | 7.983  | 537.580            | 500.000           | 7.5   |
| Aroclor-1260-4       | 8.100  | 8.008     | 8.208  | 581.670            | 500.000           | 16.3  |
| Aroclor-1260-5       | 8.421  | 8.328     | 8.528  | 508.500            | 500.000           | 1.7   |
| Decachlorobiphenyl   | 10.239 | 10.152    | 10.352 | 48.460             | 50.000            | -3.1  |
| Tetrachloro-m-xylene | 4.519  | 4.424     | 4.624  | 48.130             | 50.000            | -3.7  |

### CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1626 SAS No.: Q1626 SDG NO.: Q1626

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/24/2025

Lab Sample No.: AR1660CCC500 Data File : PP070824.D Time Analyzed: 09:06

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.906 | 4.817     | 5.017 | 512.260            | 500.000           | 2.5   |
| Aroclor-1016-2       | 4.926 | 4.836     | 5.036 | 494.660            | 500.000           | -1.1  |
| Aroclor-1016-3       | 5.103 | 5.013     | 5.213 | 496.150            | 500.000           | -0.8  |
| Aroclor-1016-4       | 5.144 | 5.054     | 5.254 | 505.180            | 500.000           | 1.0   |
| Aroclor-1016-5       | 5.359 | 5.270     | 5.470 | 508.200            | 500.000           | 1.6   |
| Aroclor-1260-1       | 6.395 | 6.307     | 6.507 | 475.200            | 500.000           | -5.0  |
| Aroclor-1260-2       | 6.583 | 6.495     | 6.695 | 462.260            | 500.000           | -7.5  |
| Aroclor-1260-3       | 6.737 | 6.649     | 6.849 | 452.510            | 500.000           | -9.5  |
| Aroclor-1260-4       | 7.208 | 7.121     | 7.321 | 438.910            | 500.000           | -12.2 |
| Aroclor-1260-5       | 7.450 | 7.362     | 7.562 | 432.640            | 500.000           | -13.5 |
| Decachlorobiphenyl   | 8.864 | 8.781     | 8.981 | 40.150             | 50.000            | -19.7 |
| Tetrachloro-m-xylene | 3.821 | 3.729     | 3.929 | 49.440             | 50.000            | -1.1  |