

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1626 SAS No.: Q1626 SDG NO.: Q1626

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL04 Date Analyzed: 03/25/2025

Lab Sample No.: AR1660CCC500 Data File : PP070854.D Time Analyzed: 14:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.672	5.577	5.777	451.340	500.000	-9.7
Aroclor-1016-2	5.694	5.599	5.799	482.170	500.000	-3.6
Aroclor-1016-3	5.756	5.661	5.861	448.330	500.000	-10.3
Aroclor-1016-4	5.853	5.758	5.958	443.420	500.000	-11.3
Aroclor-1016-5	6.146	6.051	6.251	459.890	500.000	-8.0
Aroclor-1260-1	7.263	7.171	7.371	515.030	500.000	3.0
Aroclor-1260-2	7.517	7.425	7.625	549.020	500.000	9.8
Aroclor-1260-3	7.876	7.783	7.983	537.720	500.000	7.5
Aroclor-1260-4	8.100	8.008	8.208	499.760	500.000	0.0
Aroclor-1260-5	8.421	8.328	8.528	529.700	500.000	5.9
Decachlorobiphenyl	10.239	10.152	10.352	49.010	50.000	-2.0
Tetrachloro-m-xylene	4.520	4.424	4.624	48.600	50.000	-2.8

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1626 SAS No.: Q1626 SDG NO.: Q1626

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL04 Date Analyzed: 03/25/2025

Lab Sample No.: AR1660CCC500 Data File : PP070854.D Time Analyzed: 14:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.907	4.817	5.017	511.020	500.000	2.2
Aroclor-1016-2	4.926	4.836	5.036	497.590	500.000	-0.5
Aroclor-1016-3	5.103	5.013	5.213	511.060	500.000	2.2
Aroclor-1016-4	5.145	5.054	5.254	523.950	500.000	4.8
Aroclor-1016-5	5.359	5.270	5.470	522.730	500.000	4.5
Aroclor-1260-1	6.396	6.307	6.507	491.070	500.000	-1.8
Aroclor-1260-2	6.584	6.495	6.695	470.690	500.000	-5.9
Aroclor-1260-3	6.738	6.649	6.849	470.120	500.000	-6.0
Aroclor-1260-4	7.209	7.121	7.321	440.870	500.000	-11.8
Aroclor-1260-5	7.450	7.362	7.562	424.990	500.000	-15.0
Decachlorobiphenyl	8.865	8.781	8.981	42.120	50.000	-15.8
Tetrachloro-m-xylene	3.821	3.729	3.929	51.150	50.000	2.3