

Surrogate Summary

SDG No.: Q1626

Client: Walsh Construction Company II, LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP070418.D	PIBLK-PP070418.D	Tetrachloro-m-xylene	1	20	22.0	110		60	140
		Decachlorobiphenyl	1	20	23.1	116		60	140
		Tetrachloro-m-xylene	2	20	23.0	115		60	140
		Decachlorobiphenyl	2	20	23.5	117		60	140
I.BLK-PP070828.D	PIBLK-PP070828.D	Tetrachloro-m-xylene	1	20	19.4	97		60	140
		Decachlorobiphenyl	1	20	21.0	105		60	140
		Tetrachloro-m-xylene	2	20	22.2	111		60	140
		Decachlorobiphenyl	2	20	18.1	91		60	140
PB167271BL	PB167271BL	Tetrachloro-m-xylene	1	20	21.4	107		32	144
		Decachlorobiphenyl	1	20	22.4	112		32	175
		Tetrachloro-m-xylene	2	20	23.1	115		32	144
		Decachlorobiphenyl	2	20	19.9	99		32	175
PB167271BS	PB167271BS	Tetrachloro-m-xylene	1	20	21.2	106		32	144
		Decachlorobiphenyl	1	20	22.4	112		32	175
		Tetrachloro-m-xylene	2	20	24.0	120		32	144
		Decachlorobiphenyl	2	20	19.6	98		32	175
Q1624-01MS	OK-01-03212025MS	Tetrachloro-m-xylene	1	20	22.6	113		32	144
		Decachlorobiphenyl	1	20	20.7	104		32	175
		Tetrachloro-m-xylene	2	20	23.8	119		32	144
		Decachlorobiphenyl	2	20	14.9	75		32	175
Q1624-01MSD	OK-01-03212025MSD	Tetrachloro-m-xylene	1	20	21.6	108		32	144
		Decachlorobiphenyl	1	20	19.6	98		32	175
		Tetrachloro-m-xylene	2	20	24.3	122		32	144
		Decachlorobiphenyl	2	20	17.3	86		32	175
Q1626-01	CO-32-1	Tetrachloro-m-xylene	1	20	17.0	85		32	144
		Decachlorobiphenyl	1	20	21.4	107		32	175
		Tetrachloro-m-xylene	2	20	19.6	98		32	144
		Decachlorobiphenyl	2	20	17.3	86		32	175
I.BLK-PP070839.D	PIBLK-PP070839.D	Tetrachloro-m-xylene	1	20	19.8	99		60	140
		Decachlorobiphenyl	1	20	21.3	106		60	140
		Tetrachloro-m-xylene	2	20	22.9	115		60	140
		Decachlorobiphenyl	2	20	18.1	90		60	140
I.BLK-PP070845.D	PIBLK-PP070845.D	Tetrachloro-m-xylene	1	20	19.8	99		60	140
		Decachlorobiphenyl	1	20	21.3	106		60	140
		Tetrachloro-m-xylene	2	20	24.3	122		60	140
		Decachlorobiphenyl	2	20	18.3	91		60	140
Q1626-01DL	CO-32-1DL	Tetrachloro-m-xylene	1	20	18.1	90		32	144
		Decachlorobiphenyl	1	20	22.0	110		32	175
		Tetrachloro-m-xylene	2	20	21.1	106		32	144
		Decachlorobiphenyl	2	20	17.9	90		32	175
I.BLK-PP070858.D	PIBLK-PP070858.D	Tetrachloro-m-xylene	1	20	17.2	86		60	140

Surrogate Summary

SDG No.: Q1626

Client: Walsh Construction Company II, LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP070858.D	PIBLK-PP070858.D	Decachlorobiphenyl	1	20	18.7	94		60	140
		Tetrachloro-m-xylene	2	20	18.8	94		60	140
		Decachlorobiphenyl	2	20	16.3	82		60	140