

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: WALS01

Lab Code: CHEM **Case No.:** Q1626 **SAS No.:** Q1626 **SDG NO.:** Q1626

Instrument ID: ECD_L **Calibration Date(s):** 03/11/2025 03/11/2025
Calibration Times: 10:35 11:29

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 = <u>PL094569.D</u>		CF 075 = <u>PL094570.D</u>			
CF 050 = <u>PL094571.D</u>		CF 025 = <u>PL094572.D</u>		CF 005 = <u>PL094573.D</u>			
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	2054790000	2038060000	2097220000	2238950000	2401320000	2166070000	7
4,4'-DDE	2800820000	2791230000	2844170000	3026520000	3247560000	2942060000	7
4,4'-DDT	2250760000	2258790000	2282720000	2441890000	2658540000	2378540000	7
Aldrin	3486510000	3483840000	3531750000	3767520000	4191470000	3692220000	8
alpha-BHC	4050830000	3995660000	4006320000	4207130000	4501760000	4152340000	5
alpha-Chlordane	3078250000	3067240000	3145550000	3369450000	3823700000	3296840000	10
beta-BHC	1672600000	1688180000	1742910000	1923200000	2199200000	1845220000	12
Decachlorobiphenyl	1938960000	1961100000	2020410000	2196220000	2420590000	2107460000	10
delta-BHC	3731530000	3682850000	3720060000	3932930000	4404170000	3894310000	8
Dieldrin	2998920000	2977530000	3040380000	3270050000	3703810000	3198140000	10
Endosulfan I	2856050000	2854360000	2924220000	3138080000	3578300000	3070200000	10
Endosulfan II	2509950000	2495090000	2579470000	2798020000	3191190000	2714740000	11
Endosulfan sulfate	2202890000	2228260000	2295270000	2519740000	2913940000	2432020000	12
Endrin	2585940000	2584670000	2628790000	2810860000	3250100000	2772070000	10
Endrin aldehyde	1899410000	1911570000	1995020000	2162720000	2586220000	2110990000	14
Endrin ketone	2485310000	2489220000	2504750000	2734570000	3002490000	2643270000	9
gamma-BHC (Lindane)	3886090000	3824550000	3841420000	4060770000	4338530000	3990270000	5
gamma-Chlordane	3152320000	3146400000	3207830000	3429480000	3910810000	3369370000	10
Heptachlor	3645730000	3661000000	3717820000	3965990000	4417600000	3881630000	8
Heptachlor epoxide	3127680000	3124100000	3197280000	3386170000	3890740000	3345200000	10
Methoxychlor	1109170000	1148450000	1148150000	1275940000	1303650000	1197070000	7
Tetrachloro-m-xylene	2677630000	2681900000	2722740000	2922670000	3148400000	2830670000	7

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: WALS01

Lab Code: CHEM **Case No.:** Q1626 **SAS No.:** Q1626 **SDG NO.:** Q1626

Instrument ID: ECD_L **Calibration Date(s):** 03/11/2025 03/11/2025
Calibration Times: 10:35 11:29

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 100 = <u>PL094569.D</u>		CF 075 = <u>PL094570.D</u>			
CF 050 = <u>PL094571.D</u>		CF 025 = <u>PL094572.D</u>		CF 005 = <u>PL094573.D</u>			
COMPOUND	CF 100	CF 075	CF 050	CF 025	CF 005	CF	% RSD
4,4'-DDD	3744660000	3599430000	3582800000	3559420000	3493150000	3595890000	3
4,4'-DDE	4755610000	4590230000	4589540000	4707800000	4599470000	4648530000	2
4,4'-DDT	4206410000	4061180000	4014300000	4011580000	3867070000	4032110000	3
Aldrin	5043350000	4903990000	4837240000	4815490000	4781660000	4876340000	2
alpha-BHC	5696730000	5534840000	5407190000	5382700000	4935400000	5391370000	5
alpha-Chlordane	4847910000	4714870000	4711370000	4759810000	4829890000	4772770000	1
beta-BHC	2189910000	2156390000	2172170000	2250790000	2337210000	2221290000	3
Decachlorobiphenyl	3950320000	3868520000	3903950000	4095240000	4378770000	4039360000	5
delta-BHC	5245420000	5087800000	4994500000	4942360000	4739820000	5001980000	4
Dieldrin	4991170000	4852590000	4791130000	4868920000	4755280000	4851820000	2
Endosulfan I	4455940000	4360790000	4346300000	4393610000	4387020000	4388730000	1
Endosulfan II	4346730000	4237780000	4255650000	4342220000	4458930000	4328260000	2
Endosulfan sulfate	4114360000	3978360000	4032760000	4074650000	4167060000	4073440000	2
Endrin	4434550000	4325290000	4326060000	4330520000	4401830000	4363650000	1
Endrin aldehyde	3301310000	3233140000	3279090000	3389720000	3623580000	3365370000	5
Endrin ketone	4787490000	4692680000	4707500000	4868950000	4806280000	4772580000	2
gamma-BHC (Lindane)	5372230000	5230100000	5132530000	5070750000	4891640000	5139450000	3
gamma-Chlordane	4946830000	4806290000	4770610000	4811470000	4806890000	4828420000	1
Heptachlor	5363060000	5262730000	5228770000	5273220000	5215450000	5268650000	1
Heptachlor epoxide	4619190000	4519300000	4529520000	4593130000	4631780000	4578580000	1
Methoxychlor	2089160000	2085630000	2099350000	2155560000	2175640000	2121070000	2
Tetrachloro-m-xylene	3572520000	3517330000	3502400000	3623970000	3630140000	3569270000	2