

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1626 SAS No.: Q1626 SDG NO.: Q1626

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/25/2025

Lab Sample No.: PSTDCCC050 Data File : PL094842.D Time Analyzed: 17:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.710	6.610	6.810	51.000	50.000	2.0
4,4'-DDE	6.192	6.093	6.293	49.440	50.000	-1.1
4,4'-DDT	7.024	6.924	7.124	48.070	50.000	-3.9
Aldrin	5.256	5.156	5.356	44.220	50.000	-11.6
alpha-BHC	3.994	3.894	4.094	46.490	50.000	-7.0
alpha-Chlordane	6.018	5.918	6.118	45.140	50.000	-9.7
beta-BHC	4.526	4.425	4.625	45.870	50.000	-8.3
Decachlorobiphenyl	9.055	8.956	9.156	48.190	50.000	-3.6
delta-BHC	4.773	4.673	4.873	47.330	50.000	-5.3
Dieldrin	6.345	6.244	6.444	46.960	50.000	-6.1
Endosulfan I	6.069	5.969	6.169	45.330	50.000	-9.3
Endosulfan II	6.794	6.694	6.894	46.380	50.000	-7.2
Endosulfan sulfate	7.159	7.059	7.259	46.500	50.000	-7.0
Endrin	6.572	6.474	6.674	44.340	50.000	-11.3
Endrin aldehyde	6.924	6.824	7.024	44.810	50.000	-10.4
Endrin ketone	7.645	7.544	7.744	47.010	50.000	-6.0
gamma-BHC (Lindane)	4.327	4.227	4.427	46.450	50.000	-7.1
gamma-Chlordane	5.939	5.840	6.040	44.770	50.000	-10.5
Heptachlor	4.915	4.815	5.015	44.150	50.000	-11.7
Heptachlor epoxide	5.683	5.583	5.783	43.930	50.000	-12.1
Methoxychlor	7.501	7.400	7.600	49.250	50.000	-1.5
Tetrachloro-m-xylene	3.538	3.438	3.638	47.060	50.000	-5.9

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1626 SAS No.: Q1626 SDG NO.: Q1626

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/25/2025

Lab Sample No.: PSTDCCC050 Data File : PL094842.D Time Analyzed: 17:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.778	5.682	5.882	53.380	50.000	6.8
4,4'-DDE	5.224	5.127	5.327	52.580	50.000	5.2
4,4'-DDT	6.029	5.932	6.132	51.250	50.000	2.5
Aldrin	4.219	4.122	4.322	50.840	50.000	1.7
alpha-BHC	3.272	3.174	3.374	52.990	50.000	6.0
alpha-Chlordane	5.035	4.938	5.138	50.010	50.000	0.0
beta-BHC	3.902	3.805	4.005	52.050	50.000	4.1
Decachlorobiphenyl	7.904	7.807	8.007	48.470	50.000	-3.1
delta-BHC	4.131	4.033	4.233	53.330	50.000	6.7
Dieldrin	5.355	5.258	5.458	50.480	50.000	1.0
Endosulfan I	5.091	4.994	5.194	49.970	50.000	-0.1
Endosulfan II	5.926	5.829	6.029	49.200	50.000	-1.6
Endosulfan sulfate	6.328	6.231	6.431	50.080	50.000	0.2
Endrin	5.631	5.534	5.734	49.880	50.000	-0.2
Endrin aldehyde	6.105	6.008	6.208	47.630	50.000	-4.7
Endrin ketone	6.834	6.736	6.936	48.820	50.000	-2.4
gamma-BHC (Lindane)	3.601	3.504	3.704	52.240	50.000	4.5
gamma-Chlordane	4.971	4.874	5.074	50.590	50.000	1.2
Heptachlor	3.939	3.842	4.042	49.540	50.000	-0.9
Heptachlor epoxide	4.721	4.625	4.825	50.470	50.000	0.9
Methoxychlor	6.604	6.507	6.707	51.620	50.000	3.2
Tetrachloro-m-xylene	2.769	2.672	2.872	51.850	50.000	3.7