

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 03/11/2025 03/11/2025

Client Sample No.: CCAL03 **Date Analyzed:** 03/25/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL094851.D **Time Analyzed:** 19:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.708	6.610	6.810	52.960	50.000	5.9
4,4'-DDE	6.190	6.093	6.293	52.440	50.000	4.9
4,4'-DDT	7.022	6.924	7.124	48.940	50.000	-2.1
Aldrin	5.254	5.156	5.356	46.900	50.000	-6.2
alpha-BHC	3.992	3.894	4.094	50.640	50.000	1.3
alpha-Chlordane	6.017	5.918	6.118	47.930	50.000	-4.1
beta-BHC	4.524	4.425	4.625	48.800	50.000	-2.4
Decachlorobiphenyl	9.052	8.956	9.156	50.890	50.000	1.8
delta-BHC	4.771	4.673	4.873	50.150	50.000	0.3
Dieldrin	6.342	6.244	6.444	49.720	50.000	-0.6
Endosulfan I	6.067	5.969	6.169	48.310	50.000	-3.4
Endosulfan II	6.792	6.694	6.894	48.010	50.000	-4.0
Endosulfan sulfate	7.157	7.059	7.259	47.410	50.000	-5.2
Endrin	6.570	6.474	6.674	45.150	50.000	-9.7
Endrin aldehyde	6.922	6.824	7.024	46.140	50.000	-7.7
Endrin ketone	7.642	7.544	7.744	49.210	50.000	-1.6
gamma-BHC (Lindane)	4.325	4.227	4.427	49.900	50.000	-0.2
gamma-Chlordane	5.937	5.840	6.040	47.680	50.000	-4.6
Heptachlor	4.912	4.815	5.015	46.600	50.000	-6.8
Heptachlor epoxide	5.681	5.583	5.783	47.380	50.000	-5.2
Methoxychlor	7.498	7.400	7.600	51.020	50.000	2.0
Tetrachloro-m-xylene	3.536	3.438	3.638	51.370	50.000	2.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 03/11/2025 03/11/2025

Client Sample No.: CCAL03 **Date Analyzed:** 03/25/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL094851.D **Time Analyzed:** 19:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.778	5.682	5.882	55.920	50.000	11.8
4,4'-DDE	5.223	5.127	5.327	54.930	50.000	9.9
4,4'-DDT	6.028	5.932	6.132	50.890	50.000	1.8
Aldrin	4.218	4.122	4.322	55.640	50.000	11.3
alpha-BHC	3.272	3.174	3.374	57.880	50.000	15.8
alpha-Chlordane	5.034	4.938	5.138	51.410	50.000	2.8
beta-BHC	3.902	3.805	4.005	57.070	50.000	14.1
Decachlorobiphenyl	7.903	7.807	8.007	42.010	50.000	-16.0
delta-BHC	4.130	4.033	4.233	58.310	50.000	16.6
Dieldrin	5.354	5.258	5.458	53.210	50.000	6.4
Endosulfan I	5.090	4.994	5.194	51.030	50.000	2.1
Endosulfan II	5.925	5.829	6.029	49.980	50.000	0.0
Endosulfan sulfate	6.327	6.231	6.431	49.110	50.000	-1.8
Endrin	5.630	5.534	5.734	51.350	50.000	2.7
Endrin aldehyde	6.105	6.008	6.208	46.940	50.000	-6.1
Endrin ketone	6.832	6.736	6.936	47.140	50.000	-5.7
gamma-BHC (Lindane)	3.601	3.504	3.704	57.590	50.000	15.2
gamma-Chlordane	4.971	4.874	5.074	52.030	50.000	4.1
Heptachlor	3.939	3.842	4.042	54.370	50.000	8.7
Heptachlor epoxide	4.721	4.625	4.825	53.160	50.000	6.3
Methoxychlor	6.603	6.507	6.707	51.470	50.000	2.9
Tetrachloro-m-xylene	2.770	2.672	2.872	56.870	50.000	13.7