

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 03/11/2025 03/11/2025

Client Sample No.: CCAL04 **Date Analyzed:** 03/25/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL094863.D **Time Analyzed:** 22:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.708	6.610	6.810	51.470	50.000	2.9
4,4'-DDE	6.190	6.093	6.293	48.800	50.000	-2.4
4,4'-DDT	7.021	6.924	7.124	47.260	50.000	-5.5
Aldrin	5.254	5.156	5.356	44.870	50.000	-10.3
alpha-BHC	3.992	3.894	4.094	46.930	50.000	-6.1
alpha-Chlordane	6.016	5.918	6.118	45.130	50.000	-9.7
beta-BHC	4.524	4.425	4.625	45.950	50.000	-8.1
Decachlorobiphenyl	9.051	8.956	9.156	49.060	50.000	-1.9
delta-BHC	4.771	4.673	4.873	47.600	50.000	-4.8
Dieldrin	6.342	6.244	6.444	45.510	50.000	-9.0
Endosulfan I	6.067	5.969	6.169	45.100	50.000	-9.8
Endosulfan II	6.792	6.694	6.894	44.860	50.000	-10.3
Endosulfan sulfate	7.157	7.059	7.259	45.670	50.000	-8.7
Endrin	6.571	6.474	6.674	42.730	50.000	-14.5
Endrin aldehyde	6.922	6.824	7.024	45.480	50.000	-9.0
Endrin ketone	7.643	7.544	7.744	47.520	50.000	-5.0
gamma-BHC (Lindane)	4.325	4.227	4.427	46.840	50.000	-6.3
gamma-Chlordane	5.937	5.840	6.040	45.050	50.000	-9.9
Heptachlor	4.912	4.815	5.015	44.480	50.000	-11.0
Heptachlor epoxide	5.680	5.583	5.783	45.420	50.000	-9.2
Methoxychlor	7.498	7.400	7.600	47.920	50.000	-4.2
Tetrachloro-m-xylene	3.536	3.438	3.638	47.390	50.000	-5.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** WALS01
Lab Code: ACE **SDG NO.:** Q1626
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 03/11/2025 03/11/2025

Client Sample No.: CCAL04 **Date Analyzed:** 03/25/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL094863.D **Time Analyzed:** 22:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.778	5.682	5.882	55.850	50.000	11.7
4,4'-DDE	5.221	5.127	5.327	52.650	50.000	5.3
4,4'-DDT	6.028	5.932	6.132	50.070	50.000	0.1
Aldrin	4.218	4.122	4.322	50.750	50.000	1.5
alpha-BHC	3.271	3.174	3.374	52.410	50.000	4.8
alpha-Chlordane	5.034	4.938	5.138	50.120	50.000	0.2
beta-BHC	3.902	3.805	4.005	52.030	50.000	4.1
Decachlorobiphenyl	7.903	7.807	8.007	49.180	50.000	-1.6
delta-BHC	4.130	4.033	4.233	52.790	50.000	5.6
Dieldrin	5.353	5.258	5.458	50.320	50.000	0.6
Endosulfan I	5.090	4.994	5.194	41.850	50.000	-16.3
Endosulfan II	5.925	5.829	6.029	50.060	50.000	0.1
Endosulfan sulfate	6.327	6.231	6.431	48.680	50.000	-2.6
Endrin	5.630	5.534	5.734	49.920	50.000	-0.2
Endrin aldehyde	6.104	6.008	6.208	47.550	50.000	-4.9
Endrin ketone	6.831	6.736	6.936	50.390	50.000	0.8
gamma-BHC (Lindane)	3.601	3.504	3.704	51.790	50.000	3.6
gamma-Chlordane	4.971	4.874	5.074	50.510	50.000	1.0
Heptachlor	3.939	3.842	4.042	49.530	50.000	-0.9
Heptachlor epoxide	4.721	4.625	4.825	50.350	50.000	0.7
Methoxychlor	6.603	6.507	6.707	49.960	50.000	-0.1
Tetrachloro-m-xylene	2.769	2.672	2.872	52.200	50.000	4.4