

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL031125

Review By	Abdul	Review On	3/12/2025 12:48:59 PM
Supervise By	Ankita	Supervise On	3/12/2025 2:35:44 PM
SubDirectory	PL031125	HP Acquire Method	HP Processing Method pl031125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP242583,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL094565.D	11 Mar 2025 09:41	AR\AJ	Ok
2	I.BLK	PL094566.D	11 Mar 2025 09:55	AR\AJ	Ok
3	PEM	PL094567.D	11 Mar 2025 10:08	AR\AJ	Ok,M
4	RESCHK	PL094568.D	11 Mar 2025 10:22	AR\AJ	Ok,M
5	PSTDICC100	PL094569.D	11 Mar 2025 10:35	AR\AJ	Ok,M
6	PSTDICC075	PL094570.D	11 Mar 2025 10:49	AR\AJ	Ok,M
7	PSTDICC050	PL094571.D	11 Mar 2025 11:02	AR\AJ	Ok
8	PSTDICC025	PL094572.D	11 Mar 2025 11:16	AR\AJ	Ok
9	PSTDICC005	PL094573.D	11 Mar 2025 11:29	AR\AJ	Ok,M
10	PCHLORICC1000	PL094574.D	11 Mar 2025 11:43	AR\AJ	Ok
11	PCHLORICC750	PL094575.D	11 Mar 2025 11:57	AR\AJ	Ok
12	PCHLORICC500	PL094576.D	11 Mar 2025 12:10	AR\AJ	Ok
13	PCHLORICC250	PL094577.D	11 Mar 2025 12:24	AR\AJ	Ok
14	PCHLORICC050	PL094578.D	11 Mar 2025 12:37	AR\AJ	Ok,M
15	PTOXICC1000	PL094579.D	11 Mar 2025 12:51	AR\AJ	Ok
16	PTOXICC750	PL094580.D	11 Mar 2025 13:04	AR\AJ	Ok
17	PTOXICC500	PL094581.D	11 Mar 2025 13:18	AR\AJ	Ok
18	PTOXICC250	PL094582.D	11 Mar 2025 13:31	AR\AJ	Ok,M
19	PTOXICC100	PL094583.D	11 Mar 2025 13:45	AR\AJ	Ok,M
20	PSTDICV050	PL094584.D	11 Mar 2025 13:59	AR\AJ	Ok
21	PCHLORICV500	PL094585.D	11 Mar 2025 14:26	AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL031125

Review By	Abdul	Review On	3/12/2025 12:48:59 PM
Supervise By	Ankita	Supervise On	3/12/2025 2:35:44 PM
SubDirectory	PL031125	HP Acquire Method	HP Processing Method pl031125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP242583,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PTOXICV500	PL094586.D	11 Mar 2025 14:53	AR\AJ	Ok
23	I.BLK	PL094587.D	11 Mar 2025 17:16	AR\AJ	Ok
24	PEM	PL094588.D	11 Mar 2025 17:30	AR\AJ	Ok,M
25	PSTDCCC050	PL094589.D	11 Mar 2025 17:43	AR\AJ	Ok,M
26	PB167076BL	PL094590.D	11 Mar 2025 17:57	AR\AJ	Ok
27	PB167076BS	PL094591.D	11 Mar 2025 18:11	AR\AJ	Ok,M
28	PB167076BSD	PL094592.D	11 Mar 2025 18:44	AR\AJ	Ok,M
29	Q1494-01	PL094593.D	11 Mar 2025 18:57	AR\AJ	Not Ok
30	Q1502-11	PL094594.D	11 Mar 2025 19:11	AR\AJ	Not Ok
31	Q1502-09	PL094595.D	11 Mar 2025 19:25	AR\AJ	Dilution
32	Q1502-13	PL094596.D	11 Mar 2025 19:39	AR\AJ	Not Ok
33	I.BLK	PL094597.D	11 Mar 2025 19:52	AR\AJ	Ok
34	PSTDCCC050	PL094598.D	11 Mar 2025 20:06	AR\AJ	Ok,M
35	Q1539-01	PL094599.D	11 Mar 2025 20:33	AR\AJ	Ok,M
36	Q1539-02	PL094600.D	11 Mar 2025 20:47	AR\AJ	Ok,M
37	PB167086BL	PL094601.D	11 Mar 2025 21:01	AR\AJ	Ok
38	PB167086BS	PL094602.D	11 Mar 2025 21:14	AR\AJ	Not Ok
39	PB167087BL	PL094603.D	11 Mar 2025 21:28	AR\AJ	Ok
40	PB167087BS	PL094604.D	11 Mar 2025 21:42	AR\AJ	Not Ok
41	I.BLK	PL094605.D	11 Mar 2025 21:55	AR\AJ	Ok
42	PEM	PL094606.D	11 Mar 2025 22:09	AR\AJ	Ok,M
43	PSTDCCC050	PL094607.D	11 Mar 2025 22:23	AR\AJ	Ok,M
44	PB167077BL	PL094608.D	11 Mar 2025 22:50	AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL031125

Review By	Abdul	Review On	3/12/2025 12:48:59 PM
Supervise By	Ankita	Supervise On	3/12/2025 2:35:44 PM
SubDirectory	PL031125	HP Acquire Method	HP Processing Method pl031125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP242583,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	PB167077BS	PL094609.D	11 Mar 2025 23:04	AR\AJ	Not Ok
46	Q1534-01	PL094610.D	11 Mar 2025 23:17	AR\AJ	Ok,M
47	Q1534-07	PL094611.D	11 Mar 2025 23:31	AR\AJ	Ok,M
48	Q1534-07MS	PL094612.D	11 Mar 2025 23:44	AR\AJ	Ok,M
49	Q1534-07MSD	PL094613.D	11 Mar 2025 23:58	AR\AJ	Ok,M
50	Q1534-13	PL094614.D	12 Mar 2025 00:11	AR\AJ	Dilution
51	Q1534-19	PL094615.D	12 Mar 2025 00:25	AR\AJ	Dilution
52	Q1535-01	PL094616.D	12 Mar 2025 00:39	AR\AJ	Ok,M
53	I.BLK	PL094617.D	12 Mar 2025 00:53	AR\AJ	Ok
54	PSTDCCC050	PL094618.D	12 Mar 2025 01:06	AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL032525

Review By	Abdul	Review On	3/26/2025 11:23:53 AM
Supervise By		Supervise On	
SubDirectory	PL032525	HP Acquire Method	HP Processing Method pl031125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP242583,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PL094829.D	25 Mar 2025 10:05	AR\AJ	Ok
2	I.BLK	PL094830.D	25 Mar 2025 10:19	AR\AJ	Ok
3	PEM	PL094831.D	25 Mar 2025 11:36	AR\AJ	Ok,NS
4	PSTDCCC050	PL094832.D	25 Mar 2025 11:50	AR\AJ	Ok,NS
5	PB167272BL	PL094833.D	25 Mar 2025 12:25	AR\AJ	Ok
6	PB167272BS	PL094834.D	25 Mar 2025 14:00	AR\AJ	Ok,NS
7	Q1624-01	PL094835.D	25 Mar 2025 14:14	AR\AJ	Ok,NS
8	Q1626-01	PL094836.D	25 Mar 2025 14:27	AR\AJ	Ok,NS
9	Q1626-01MS	PL094837.D	25 Mar 2025 14:41	AR\AJ	Ok,NS
10	Q1626-01MSD	PL094838.D	25 Mar 2025 14:55	AR\AJ	Ok,NS
11	PB167290BL	PL094839.D	25 Mar 2025 15:08	AR\AJ	Ok
12	PB167290BS	PL094840.D	25 Mar 2025 15:22	AR\AJ	Ok,NS
13	I.BLK	PL094841.D	25 Mar 2025 16:15	AR\AJ	Ok
14	PSTDCCC050	PL094842.D	25 Mar 2025 17:01	AR\AJ	Ok,NS
15	Q1630-01	PL094843.D	25 Mar 2025 17:15	AR\AJ	Ok,NS
16	Q1630-01MS	PL094844.D	25 Mar 2025 17:28	AR\AJ	Ok,NS
17	Q1630-01MSD	PL094845.D	25 Mar 2025 17:42	AR\AJ	Ok,NS
18	Q1634-01	PL094846.D	25 Mar 2025 17:56	AR\AJ	Ok,NS
19	Q1634-03	PL094847.D	25 Mar 2025 18:09	AR\AJ	Ok,NS
20	Q1635-01	PL094848.D	25 Mar 2025 18:23	AR\AJ	Ok
21	I.BLK	PL094849.D	25 Mar 2025 18:37	AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL032525

Review By	Abdul	Review On	3/26/2025 11:23:53 AM
Supervise By		Supervise On	
SubDirectory	PL032525	HP Acquire Method	HP Processing Method pl031125 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23793,PP24095		
Initial Calibration Stds	PP24260,PP24261,PP24262,PP24269,PP24266,PP24267,PP24268,PP24269,PP24270,PP24271,PP24272,PP24273,PP24274,PP24275,PP24277,PP24278,PP24279,PP24280,PP24281,PP24282,PP242583,PP24284		
CCC	PP24261,PP24273,PP24279,PP24284		
Internal Standard/PEM			
ICV/I.BLK	PP24273,PP24279,PP24284		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	PEM	PL094850.D	25 Mar 2025 18:50	AR\AJ	Ok,NS
23	PSTDCCC050	PL094851.D	25 Mar 2025 19:18	AR\AJ	Ok,NS
24	PB167311BL	PL094852.D	25 Mar 2025 19:31	AR\AJ	Ok
25	PB167311BS	PL094853.D	25 Mar 2025 19:45	AR\AJ	Not Ok
26	PB167233TB	PL094854.D	25 Mar 2025 19:58	AR\AJ	Ok
27	PB167275TB	PL094855.D	25 Mar 2025 20:12	AR\AJ	Ok
28	Q1609-03	PL094856.D	25 Mar 2025 20:26	AR\AJ	Ok
29	Q1609-03MS	PL094857.D	25 Mar 2025 20:39	AR\AJ	Ok
30	Q1609-03MSD	PL094858.D	25 Mar 2025 20:53	AR\AJ	Ok,NS
31	Q1626-03	PL094859.D	25 Mar 2025 21:07	AR\AJ	Ok,NS
32	Q1627-01	PL094860.D	25 Mar 2025 21:20	AR\AJ	Ok,NS
33	Q1632-02	PL094861.D	25 Mar 2025 21:34	AR\AJ	Ok,NS
34	I.BLK	PL094862.D	25 Mar 2025 21:48	AR\AJ	Ok
35	PSTDCCC050	PL094863.D	25 Mar 2025 22:01	AR\AJ	Ok,NS

M : Manual Integration