

Surrogate Summary

SDG No.: Q1626

Client: Walsh Construction Company II, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL094566.D	PIBLK-PL094566.D	Decachlorobiphenyl	1	20	22.7	114		43	140
		Tetrachloro-m-xylene	1	20	20.6	103		77	126
		Decachlorobiphenyl	2	20	21.0	105		43	140
		Tetrachloro-m-xylene	2	20	20.2	101		77	126
I.BLK-PL094830.D	PIBLK-PL094830.D	Decachlorobiphenyl	1	20	20.2	101		43	140
		Tetrachloro-m-xylene	1	20	17.4	87		77	126
		Decachlorobiphenyl	2	20	18.5	92		43	140
		Tetrachloro-m-xylene	2	20	17.7	89		77	126
PB167272BL	PB167272BL	Decachlorobiphenyl	1	20	21.0	105		20	144
		Tetrachloro-m-xylene	1	20	17.9	90		19	148
		Decachlorobiphenyl	2	20	19.7	98		20	144
		Tetrachloro-m-xylene	2	20	19.1	95		19	148
PB167272BS	PB167272BS	Decachlorobiphenyl	1	20	21.6	108		20	144
		Tetrachloro-m-xylene	1	20	17.3	87		19	148
		Decachlorobiphenyl	2	20	20.4	102		20	144
		Tetrachloro-m-xylene	2	20	18.4	92		19	148
Q1626-01	CO-32-1	Decachlorobiphenyl	1	20	28.5	143		20	144
		Tetrachloro-m-xylene	1	20	15.5	78		19	148
		Decachlorobiphenyl	2	20	21.3	106		20	144
		Tetrachloro-m-xylene	2	20	17.5	88		19	148
Q1626-01MS	CO-32-1MS	Decachlorobiphenyl	1	20	28.5	143		20	144
		Tetrachloro-m-xylene	1	20	16.0	80		19	148
		Decachlorobiphenyl	2	20	18.8	94		20	144
		Tetrachloro-m-xylene	2	20	18.5	93		19	148
Q1626-01MSD	CO-32-1MSD	Decachlorobiphenyl	1	20	26.2	131		20	144
		Tetrachloro-m-xylene	1	20	16.0	80		19	148
		Decachlorobiphenyl	2	20	18.6	93		20	144
		Tetrachloro-m-xylene	2	20	17.6	88		19	148
I.BLK-PL094841.D	PIBLK-PL094841.D	Decachlorobiphenyl	1	20	19.4	97		43	140
		Tetrachloro-m-xylene	1	20	17.0	85		77	126
		Decachlorobiphenyl	2	20	17.2	86		43	140
		Tetrachloro-m-xylene	2	20	17.5	87		77	126